

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ac-9 Repair Summary

Date 10/13/2016 Customer University of Maine

RO # 31754

S/N # ac90274 Technician cad/dw/JAM

Cal # 11

Incoming Status/Noted Problems with ac Meter

- Customer requests no service to instrument other than extended temp run.

Standard Service

- Replaced desiccant, performed vacuum purge
- Performed temperature, air, and water calibrations
- Updated meter's calibration information
- Performed complete ac digital calibration
- Compiled customer CD with calibration, device file, and air tracking file

Additional Repairs

- Extended external temperature calibration range processed from 45 to ~1 celcius.

Comments