PO Box 518 620 Applegate St. Philomath, OR 97370



ac-s Calibration

Date 6/2/2022			
S/N # ACS-003	Fechnician cad		Cal # 18
Wavelength range:	400 to 730 nm		
Tcal:	22.4 deg C		
Ical:	19.7 deg C		
Acal:	22.7 deg C		
Temperature range:	1.3 to	35.0 deg C	
Zero:	25.5 deg C		

Definitions

Wavelength range: Wavelengths for which instrument use is recommended.

Tcal: Water temperature in flow tubes during water calibration.

Ical: Internal instrument temperature during water calibration.

Acal: Internal instrument temperature when recording air offsets in the *.cal file.

Temperature range: Recommended internal instrument operating range.

Zero: Internal instrument temperature at which the temperature correction is zero in the device file.