

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ac-s Calibration

Date 12/15/2017

S/N # acs298

Cal # 1

Wavelength range: 400 to 730 nm

Tcal: 17.8 deg C

Ical: 18.7 deg C

Acal: 23.1 deg C

Temperature range: 0.9 to 35.0 deg C

Zero: 25.5 deg C

Definitions

Wavelength range: Wavelengths for which instrument use is recommended.

Tcal: Water temperature in flow tubes during water calibration.

Ical: Internal instrument temperature during water calibration.

Acal: Internal instrument temperature when recording air offsets in the *.cal file.

Temperature range: Recommended internal instrument operating range.

Zero: Internal instrument temperature at which the temperature correction is zero in the device file.