## Temperature Calibration Report

Customer:	Woods Hole Oceangraphic Institution			
Job Number:	69924	Date of Repo	ort:	6/28/2012
Model Number	SBE 03Plus	Serial Numb	er:	03P4230
Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.  An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition				
~ .	ations (consult the SE.	he chosen coefficients. The coefficient 'o ASOFT manual). Calibration coefficient		•
Date: 6/28/2012		Drift since last cal:	+0.00012	Degrees Celsius/year
Comments:				
'CALIBRATION Date:	AFTER REPAIR'	☐ Per	rformed	✓ Not Performed  Degrees Celsius/year
Comments:	_			<u>.</u>