

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2147
CALIBRATION DATE: 29-Jun-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.00476180e+001
h = 1.40389245e+000
i = -2.77152621e-003
j = 2.56554213e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 9.21896126e-009
b = 1.39547151e+000
c = -1.00265764e+001
d = -7.05501414e-005
m = 8.0
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.68059	0.00000	0.00000
-0.9999	34.7546	2.80008	5.21987	2.80007	-0.00001
1.0000	34.7551	2.97123	5.33602	2.97124	0.00001
15.0000	34.7569	4.26510	6.14299	4.26509	-0.00000
18.5001	34.7567	4.61133	6.34142	4.61134	0.00001
29.0001	34.7561	5.69358	6.92471	5.69356	-0.00002
32.5001	34.7497	6.06570	7.11411	6.06571	0.00001


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 08-Dec-10 0.9999590
▲ 29-Jun-12 1.0000000