

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2989
CALIBRATION DATE: 03-Jul-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.03021098e+001
h = 1.39831805e+000
i = 5.87483173e-004
j = 3.66112815e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 4.44981611e-004
b = 1.39857798e+000
c = -1.03023064e+001
d = -8.36705723e-005
m = 3.3
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.71251	0.00000	0.00000
-0.9986	34.6588	2.79319	5.22146	2.79321	0.00002
1.0000	34.6593	2.96382	5.33647	2.96380	-0.00002
15.0000	34.6606	4.25453	6.13683	4.25452	-0.00001
18.5000	34.6605	4.59993	6.33377	4.59992	-0.00000
29.0000	34.6584	5.67936	6.91294	5.67940	0.00004
32.5001	34.6513	6.05047	7.10102	6.05045	-0.00003


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 20-Apr-11 0.9999893
▲ 03-Jul-12 1.0000000