

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3061
CALIBRATION DATE: 07-Jul-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -9.81107499e+000
h = 1.16948339e+000
i = 7.34483942e-005
j = 4.42963869e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 9.95036662e-005
b = 1.16941058e+000
c = -9.81061596e+000
d = -8.28932565e-005
m = 3.7
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.89570	0.00000	0.00000
-0.9999	34.9530	2.81457	5.69248	2.81454	-0.00003
1.0001	34.9529	2.98653	5.81987	2.98656	0.00003
15.0001	34.9547	4.28680	6.70455	4.28678	-0.00001
18.5001	34.9542	4.63469	6.92202	4.63473	0.00003
29.0001	34.9503	5.72180	7.56097	5.72174	-0.00005
32.5001	34.9372	6.09469	7.76802	6.09473	0.00003


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 20-Apr-11 0.9999765
▲ 07-Jul-12 1.0000000