PO Box 518 620 Applegate St. Philomath, OR 97370



C-Star Calibration

| Date | May 22, 2014 | S/N# | CST-758DR | Pathlength 25cm |
|--|--------------|------|---------------|-----------------|
| | | | Analog output | |
| V _d | | | 0.058 V | |
| | | | | |
| V_{air} | | | 4.789 V | |
| V _{ref} | | | 4.681 V | |
| | | | | |
| Temperature of calibration water | | | | 23.0 °C |
| Ambient temperature during calibration | | | | 22.1 °C |

Relationship of transmittance (Tr) to beam attenuation coefficient (c), and pathlength (x, in meters): $Tr = e^{-cx}$

To determine beam transmittance: Tr = (V_{sig} - V_{dark}) / (V_{ref} - V_{dark})

To determine beam attenuation coefficient: **c** = -1/x * In (Tr)

 V_d Meter output with the beam blocked. This is the offset.

V_{air} Meter output in air with a clear beam path.

V_{ref} Meter output with clean water in the path.

Temperature of calibration water: temperature of clean water used to obtain V_{ref}.

Ambient temperature: meter temperature in air during the calibration.

V_{sig} Measured signal output of meter.