

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2645
CALIBRATION DATE: 03-Jul-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.02340700e+001
h = 1.41384268e+000
i = -1.16258107e-003
j = 1.66811839e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 6.30663110e-006
b = 1.41108148e+000
c = -1.02290064e+001
d = -8.61141699e-005
m = 5.1
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.69227	0.00000	0.00000
-0.9986	34.6588	2.79319	5.19845	2.79320	0.00001
1.0000	34.6593	2.96382	5.31326	2.96381	-0.00001
15.0000	34.6606	4.25453	6.11187	4.25452	-0.00001
18.5000	34.6605	4.59993	6.30832	4.59992	-0.00001
29.0000	34.6584	5.67936	6.88589	5.67940	0.00004
32.5001	34.6513	6.05047	7.07340	6.05045	-0.00003

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

