

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2824
CALIBRATION DATE: 14-Feb-12

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.01362591e+001
h = 1.40333620e+000
i = 2.31373323e-004
j = 6.74743168e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 1.67923777e-004
b = 1.40361989e+000
c = -1.01368567e+001
d = -8.61927009e-005
m = 3.7
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.68650	0.00000	0.00000
-0.9999	34.9847	2.81688	5.21886	2.81688	-0.00000
1.0001	34.9849	2.98901	5.33463	2.98900	-0.00001
15.0001	34.9849	4.29011	6.13912	4.29011	0.00000
18.5001	34.9845	4.63827	6.33700	4.63832	0.00004
29.0001	34.9824	5.72646	6.91863	5.72637	-0.00009
32.5001	34.9712	6.09995	7.10729	6.10001	0.00006


Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction



● 20-Apr-11 0.9999215
▲ 14-Feb-12 1.0000000