

Temperature Calibration Report

| Customer: | Woods Hole Oceangraphic Institution | | | |
|--|-------------------------------------|-----------------------|----------|----------------------|
| Job Number: | 88327 | Date of Rep | ort: | 1/20/2016 |
| Model Number | SBE 03Plus | Serial Numl | ber: | 03P4303 |
| Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request. An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users | | | | |
| must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data. 'AS RECEIVED CALIBRATION' Performed Not Performed | | | | |
| Date: 1/20/2016 | | Drift since last cal: | -0.00026 | Degrees Celsius/year |
| Comments: | | | | |
| 'CALIBRATION | AFTER REPAIR' | - | erformed | ✓ Not Performed |
| Date: | J | Drift since Last cal: | | Degrees Celsius/year |
| Comments: | | | | |