



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Service

Report

RMA Number

77878

Customer Information:

Company

Woods Hole Oceanographic Institution

Date

2/4/2014

Contact

Robbie Laird

PO Number

TBD

Serial Number

4554695-0288

Model Number

SBE 45

Services Requested:

1. Evaluate/Repair Instrumentation.
2. Perform Routine Calibration Service.

Problems Found:

Services Performed:

1. Performed initial diagnostic evaluation.
2. Performed "Post Cruise" calibration of the temperature & conductivity sensors.
3. Cleaned and replatinized the conductivity cell.
4. Performed "Final" calibration of the temperature & conductivity sensors.
5. Performed internal inspection and O-ring replacement.
6. Performed complete system check and full diagnostic evaluation.

Special Notes: