



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th St. Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer Woods Hole Oceanographic Institution
Job Number 71602
Date 2/13/2013
Technician AL

Serial Number 480019

Low Pressure (PSI) 45 PSI
Time (Minutes) 15 Minutes

High Pressure (PSI) 45 PSI
Time (Minutes) 15 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings, again. *The unit is not designed for high pressure applications.

