

Sea-Bird Electronics, Inc.

13431 NE 20th Street, Bellevue, WA 98005-2010 USA

Phone: (+1) 425-643-9866 Fax (+1) 425-643-9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 2880
CALIBRATION DATE: 10-Mar-17

SBE 4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.06660155e+001
h = 1.44280780e+000
i = -2.68565413e-005
j = 7.52174506e-005

CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.71847	0.00000	0.00000
-1.0000	34.6149	2.78986	5.16663	2.78984	-0.00002
1.0000	34.6151	2.96040	5.27952	2.96042	0.00002
15.0000	34.6155	4.24958	6.06473	4.24956	-0.00002
18.5000	34.6137	4.59438	6.25800	4.59440	0.00002
29.0001	34.6024	5.67122	6.82611	5.67122	-0.00001
32.5001	34.5872	6.04055	7.01029	6.04056	0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

