

SBE 41CP CERTIFICATES

CTD Serial Number 41CP-3818

Instrument Configuration.....	1
DC - Coefficient Output File.....	2
Temperature Calibration Sheet.....	3
Conductivity Calibration Sheet.....	4
Pressure Calibration Sheet.....	5

SEA-BIRD ELECTRONICS, INC.
1808 136th Place NE
Bellevue, Washington 98005 USA
Phone: (425) 643 9866
Fax: (425) 643 9954
Email: seabird@seabird.com

SBE 41CP Instrument Configuration

Model Number: SBE 41CP

Serial Number: 41CP-3818

Part Number: 90365.051

Description : APEX SBE 41CP Configuration

Firmware Version: 2.0

Pressure Type: Druck

Pressure Range: 2000 Dbar

Pressure Serial Number: 3198523

SBE 41CP UW V 2.0 SERIAL NO. 3818

temperature: 20-may-11

TA0 = 1.394506e-05

TA1 = 2.697308e-04

TA2 = -2.140708e-06

TA3 = 1.398875e-07

conductivity: 20-may-11

G = -9.981592e-01

H = 1.528333e-01

I = -4.142948e-04

J = 5.271134e-05

CPCOR = -9.570001e-08

CTCOR = 3.250000e-06

WBOTC = 9.107778e-08

pressure S/N = 3198523, range = 2900 psia: 12-may-11

PA0 = 1.583229e-01

PA1 = 1.396848e-01

PA2 = -4.058034e-08

PTCA0 = 2.147064e+01

PTCA1 = 1.961437e-01

PTCA2 = -6.675687e-03

PTCB0 = 2.528250e+01

PTCB1 = 3.000000e-04

PTCB2 = 0.000000e+00

PTHA0 = -7.655653e+01

PTHA1 = 4.937304e-02

PTHA2 = -4.202479e-07

POFFSET = 0.000000e+00

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3818
CALIBRATION DATE: 20-May-11

SBE 41_{CP} TEMPERATURE CALIBRATION DATA
ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = 1.394506e-005

a1 = 2.697308e-004


a2 = -2.140708e-006

a3 = 1.398875e-007

BATH TEMP (ITS-90)	INSTRUMENT OUTPUT	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	832810.0	1.0000	0.0000
4.5000	710230.5	4.4999	-0.0001
15.0000	449319.6	15.0000	0.0000
18.5000	388155.6	18.5000	0.0000
23.9940	310356.6	23.9940	-0.0000
29.0000	254692.4	28.9999	-0.0001
32.5000	222562.2	32.5000	0.0000

Temperature ITS-90 = $1/\{a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3818
CALIBRATION DATE: 20-May-11

SBE 41_{CP} CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.981592e-001

CPcor = -9.5700e-008

h = 1.528333e-001

CTcor = 3.2500e-006

i = -4.142948e-004

WBOTC = 9.1078e-008

j = 5.271134e-005


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2561.59	0.00000	0.00000
1.0000	34.9242	2.98431	5117.07	2.98431	0.00000
4.5000	34.9043	3.29221	5310.76	3.29220	-0.00000
15.0000	34.8616	4.27658	5886.69	4.27658	-0.00000
18.5000	34.8523	4.62263	6075.94	4.62263	0.00000
23.9940	34.8420	5.18140	6369.36	5.18140	0.00000
29.0000	34.8357	5.70514	6632.28	5.70514	0.00000
32.5000	34.8315	6.07834	6813.24	6.07834	-0.00000

$f = \text{INST FREQ} * \sqrt{1.0 + \text{WBOTC} * t} / 1000.0$

Conductivity = $(g + hf^2 + if^3 + jf^4) / (1 + \delta t + \epsilon p)$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = instrument conductivity - bath conductivity



SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3818

CALIBRATION DATE: 12-May-11

SBE 41cp PRESSURE CALIBRATION DATA

2900 psia S/N 3198523

COEFFICIENTS:

PA0 = 1.583229e-001

PA1 = 1.396848e-001

PA2 = -4.058034e-008

PTHA0 = -7.655653e+001

PTHA1 = 4.937304e-002

PTHA2 = -4.202479e-007

PTCA0 = 2.147064e+001

PTCA1 = 1.961437e-001

PTCA2 = -6.675687e-003

PTCB0 = 2.528250e+001

PTCB1 = 3.000000e-004

PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	THERMISTOR OUTPUT	COMPUTED PRESSURE	ERROR %FSR
14.72	127.1	2020.2	14.75	0.00
591.80	4263.4	2022.9	591.65	-0.01
1168.63	8411.6	2025.1	1168.82	0.01
1745.53	12567.3	2026.0	1745.63	0.00
2322.39	16732.9	2027.6	2322.41	0.00
2898.97	20907.2	2028.4	2898.98	0.00
2322.43	16732.4	2028.3	2322.34	-0.00
1745.84	12568.6	2028.3	1745.81	-0.00
1168.69	8410.3	2028.5	1168.64	-0.00
591.84	4264.0	2028.2	591.74	-0.00
14.71	127.2	2037.8	14.78	0.00

THERMAL CORRECTION

TEMP ITS90	PRESS TEMP	INST OUTPUT
32.50	2251.70	128.96
29.00	2178.80	128.40
23.99	2073.00	129.34
18.50	1958.00	130.55
15.00	1884.40	130.78
4.50	1665.40	129.97
1.00	1592.40	128.88

TEMP(ITS90)	SPAN(mV)
-5.00	25.28
35.00	25.29

$$y = \text{thermistor output}; t = \text{PTHA0} + \text{PTHA1} * y + \text{PTHA2} * y^2$$

$$x = \text{pressure output} - \text{PTCA0} - \text{PTCA1} * t - \text{PTCA2} * t^2$$

$$n = x * \text{PTCB0} / (\text{PTCB0} + \text{PTCB1} * t + \text{PTCB2} * t^2)$$

$$\text{pressure (psia)} = \text{PA0} + \text{PA1} * n + \text{PA2} * n^2$$

Date, Avg Delta P %FS

12-May-11 -0.00

