



SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Temperature Calibration Report

Customer:	EMS/France		
Job Number:	55354	Date of Report:	7/31/2009
Model Number:	SBE 03Plus	Serial Number:	03P2492

Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients using the program SEACON. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.

'AS RECEIVED CALIBRATION'

Performed Not Performed

Date: 7/17/2009

Drift since last cal: +0.00262 Degrees Celsius/year

Comments:

'FINAL CALIBRATION'

Performed Not Performed

Date: 7/31/2009

Drift since 10 Jun 08 +0.00268 Degrees Celsius/year

Comments:

Added wet-pluggable connector.