



Sea-Bird GmbH
Postfach 1167, 87401 Kempten, Germany
Phone: +49 831 9 60994 701 Fax: +49 831 960994 709
Email: seabird.eu@seabird.com

Temperature Calibration Report

Customer:	CNRS/LOV		
Job Number:	E00302	Date of Report:	12/19/2013
Model Number:	SBE 03Plus	Serial Number:	03P2492

Temperature sensors are normally calibrated 'as received', without adjustments, allowing a determination sensor drift. If the calibration identifies a problem, then a second calibration is performed after work is completed. The 'as received' calibration is not performed if the sensor is damaged or non-functional, or by customer request.

An 'as received' calibration certificate is provided, listing coefficients to convert sensor frequency to temperature. Users must choose whether the 'as received' calibration or the previous calibration better represents the sensor condition during deployment. In SEASOFT enter the chosen coefficients. The coefficient 'offset' allows a small correction for drift between calibrations (consult the SEASOFT manual). Calibration coefficients obtained after a repair apply only to subsequent data.

'AS RECEIVED CALIBRATION'

Performed **Not Performed**

Date: 12/19/2013

Drift since last cal: +0.00038 Degrees Celsius/year

Comments:

'CALIBRATION AFTER REPAIR'

Performed **Not Performed**

Date:

Drift since Last cal: Degrees Celsius/year

Comments: