

SEA-BIRD ELECTRONICS, INC.

13431 NE 20th Street, Bellevue, Washington, 98005-2010 USA

Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 3598
CALIBRATION DATE: 02-Sep-11

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.00103102e+001
h = 1.54487792e+000
i = -2.41739215e-003
j = 2.77547890e-004
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 6.01730038e-007
b = 1.53865836e+000
c = -9.99840846e+000
d = -8.59113593e-005
m = 6.3
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.54912	0.00000	0.00000
-1.0000	34.6414	2.79180	4.96318	2.79178	-0.00002
1.0000	34.6419	2.96247	5.07358	2.96249	0.00001
15.0000	34.6415	4.25243	5.84037	4.25246	0.00003
18.5000	34.6416	4.59769	6.02890	4.59766	-0.00003
29.0000	34.6393	5.67658	6.58306	5.67658	-0.00000
32.5000	34.6306	6.04726	6.76284	6.04726	0.00000

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

