



Sea-Bird GmbH
Postfach 1167
87401 Kempten
Germany

+49 831 9 60994 701
seabird.eu@seabird.com
www.seabird.com

Pressure Test Certificate

Test Date: **2024-02-14**

Description: **SBE-45 Mini Thermosalinograph**

Sensor Information:

Replaced the main piston "O"-Rings.

*The unit is not designed for high pressure applications.

Model Number: **SBE-45**

Serial Number: **0286**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** Minutes

High Pressure Test: **40** PSI Held For: **30** Minutes

Passed Test: **True**

Tested By: **SW**

High pressure is generally equal to the maximum depth rating of the instrument

