



Sea-Bird GmbH  
 Postfach 1167  
 87401 Kempten  
 Germany

+49 831 9 60994 701  
 seabird.eu@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 3598  
 CALIBRATION DATE: 16-Jan-24

SBE 4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00093251e+001  
 h = 1.54430337e+000  
 i = -2.22431335e-003  
 j = 2.60524359e-004

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.54916	0.00000	0.00000
-1.0000	34.8035	2.80364	4.97092	2.80365	0.00000
1.0000	34.8045	2.97505	5.08159	2.97503	-0.00002
15.0000	34.8043	4.27030	5.85029	4.27036	0.00006
18.5000	34.8037	4.61688	6.03922	4.61686	-0.00002
29.0000	34.7981	5.69967	6.59441	5.69960	-0.00007
32.5001	34.7856	6.07125	6.77438	6.07130	0.00005

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = (g + h \* f<sup>2</sup> + i \* f<sup>3</sup> + j \* f<sup>4</sup>) / 10 (1 + δ \* t + ε \* p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

