



Sea-Bird GmbH  
 Postfach 1167  
 87401 Kempten  
 Germany

+49 831 9 60994 701  
 seabird.eu@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 0286  
 CALIBRATION DATE: 13-Feb-24

SBE 45 TEMPERATURE CALIBRATION DATA  
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

a0 = 1.770314e-005  
 a1 = 2.846229e-004  
 a2 = -3.215254e-006  
 a3 = 1.728565e-007

BATH TEMP (° C)	INSTRUMENT OUTPUT (counts)	INST TEMP (° C)	RESIDUAL (° C)
1.0000	610749.2	1.0000	-0.0000
4.5000	521312.8	4.5001	0.0001
15.0000	330634.7	14.9999	-0.0001
18.5000	285852.5	18.5000	-0.0000
24.0000	228779.8	24.0001	0.0001
29.0000	187994.6	29.0000	-0.0000
32.5000	164400.4	32.5000	0.0000

n = Instrument Output (counts)

$$\text{Temperature ITS-90 (°C)} = 1 / \{ a_0 + a_1[\ln(n)] + a_2[\ln^2(n)] + a_3[\ln^3(n)] \} - 273.15$$

Residual (°C) = instrument temperature - bath temperature

