


PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

---

## C-Star Repairs

Date 1/17/2020 Customer Fondation Tara Ocean  
S/N# **CST-1269DR** Repair Order 1005509705

---

### Standard Service (Analog)

- Replaced cover plate seal
- Checked optics alignment
- Performed noise test: 1 sample/sec for 60 sec
- Performed stability test: 1 sample/min for 12 hrs
- Performed temperature test
- Performed water calibration
- Shake-tested unit
- Pressure-tested unit
- Updated unit's calibration sheet

### Additional Repairs

None

### Comments

Performed a post deployment calibration at the request of the customer