



Sea-Bird GmbH  
 Postfach 1167  
 87401 Kempten  
 Germany

+49 831 9 60994 701  
 seabird.eu@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 3679  
 CALIBRATION DATE: 17-Dec-19

SBE 4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00999235e+001  
 h = 1.52245788e+000  
 i = 8.95664311e-004  
 j = 3.72188922e-005

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.57349	0.00000	0.00000
-1.0000	34.8285	2.80547	4.99724	2.80546	-0.00001
1.0000	34.8293	2.97697	5.10810	2.97697	0.00000
15.0000	34.8303	4.27315	5.87827	4.27315	0.00000
18.5000	34.8295	4.61993	6.06766	4.61994	0.00001
29.0000	34.8266	5.70382	6.62449	5.70380	-0.00001
32.5000	34.8129	6.07546	6.80486	6.07547	0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars);  $\delta$  = CTcor;  $\epsilon$  = CPcor;

Conductivity (S/m) =  $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

