



**SEA-BIRD ELECTRONICS, INC.**  
1808 - 136th Place Northeast, Bellevue, Washington 98005 USA  
Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

**Customer** EMS/France

**Job Number** 55354

**Date** 8/12/2009

**Technician** AS

<b><u>Serial Number</u></b>	17P26075-0256
-----------------------------	---------------

**Low Pressure (PSI)** 45 PSI

**Time (Minutes)** 15 Minutes

**High Pressure (PSI)** 10000 PSI

**Time (Minutes)** 45 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. Upgraded to wet pluggable connectors.

