

## Pressure Test Certificate

Test Date: **2021-06-17**Description: **SBE-3 Temperature Sensor**

### Sensor Information:

Model Number: **SBE-3**Serial Number: **5147**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      MinutesHigh Pressure Test: **8000**      PSI      Held For: **30**      MinutesPassed Test: **True**Tested By: **SW**

**High pressure is generally equal to the maximum depth rating of the instrument**

