



Sea-Bird GmbH
 Postfach 1167
 87401 Kempten
 Germany

+49 831 9 60994 701
 seabird.eu@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 2492
 CALIBRATION DATE: 17-Dec-19

SBE 3 TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

COEFFICIENTS:

g = 4.40031588e-003
 h = 6.61585436e-004
 i = 2.59798904e-005
 j = 2.48755666e-006
 f0 = 1000.0

BATH TEMP (° C)	INSTRUMENT OUTPUT (Hz)	INST TEMP (° C)	RESIDUAL (° C)
-1.5000	3101.230	-1.5000	0.00001
1.0000	3276.271	1.0000	0.00001
4.5000	3533.146	4.5000	-0.00004
8.0000	3804.154	8.0000	-0.00003
11.5000	4089.657	11.5000	0.00001
15.0000	4390.005	15.0000	0.00005
18.5000	4705.535	18.5000	0.00004
22.0000	5036.572	22.0000	-0.00003
25.5000	5383.446	25.5000	-0.00001
29.0000	5746.443	28.9999	-0.00007
32.5000	6125.878	32.5001	0.00006

f = Instrument Output (Hz)

$$\text{Temperature ITS-90 (°C)} = 1 / \{g + h[\ln(f_0 / f)] + i[\ln^2(f_0 / f)] + j[\ln^3(f_0 / f)]\} - 273.15$$

$$\text{Residual (°C)} = \text{instrument temperature} - \text{bath temperature}$$

