

Pressure Test Certificate

Test Date: **2019-12-18**Description: **SBE-3 Temperature Sensor**

Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SBE-3**Serial Number: **5122**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** MinutesHigh Pressure Test: **8000** PSI Held For: **15** MinutesPassed Test: **True**Tested By: **SH**

High pressure is generally equal to the maximum depth rating of the instrument

