

SEA-BIRD ELECTRONICS, INC. 1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

1808 - 136th Place Northeast, Bellevue, Washington 98005 USAPhone: (425) 643-9866Fax: (425) 643-9954Www.seabird.com

Pressure Test Certificate

<u>Customer</u>	EMS/France
<u>Job Number</u>	55354
<u>Date</u>	7/22/2009
<u>Technician</u>	BLT
Serial Number	042039
Low Pressure (PSI)	50 PSI
<u>Time (Minutes)</u>	15 Minutes
<u>High Pressure (PSI)</u> <u>Time (Minutes)</u>	2 10000 PSI 30 Minutes
Pass 🗹 Fail 🗌	
Comments	
Replaced the main	piston "O"-Rings. Replaced 3-pin connector with 3 pin wet-pluggable.
L	

