

Sea-Bird GmbH

Postfach 1167, 87401 Kempten, Germany

Phone: +49 831 960994 701 Fax: +49 831 960994 709 Email: seabird.eu@seabird.com

SENSOR SERIAL NUMBER: 2039
CALIBRATION DATE: 03-Apr-13

SBE4 CONDUCTIVITY CALIBRATION DATA
PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

GHIJ COEFFICIENTS

g = -1.03753564e+001
h = 1.27194655e+000
i = -2.94075536e-004
j = 8.22234174e-005
CPcor = -9.5700e-008 (nominal)
CTcor = 3.2500e-006 (nominal)

ABCDM COEFFICIENTS

a = 3.60906496e-005
b = 1.27123413e+000
c = -1.03738712e+001
d = -8.33462434e-005
m = 4.2
CPcor = -9.5700e-008 (nominal)

BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (kHz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
0.0000	0.0000	0.00000	2.85625	0.00000	0.00000
-1.0000	34.7930	2.80287	5.49298	2.80290	0.00003
1.0000	34.7940	2.97424	5.61400	2.97421	-0.00003
15.0000	34.7927	4.26902	6.45536	4.26904	0.00002
18.5000	34.7915	4.61543	6.66232	4.61541	-0.00002
28.9999	34.7822	5.69735	7.27063	5.69737	0.00001
32.5000	34.7620	6.06759	7.46730	6.06758	-0.00001

Conductivity = $(g + hf^2 + if^3 + jf^4) / 10(1 + \delta t + \epsilon p)$ Siemens/meter

Conductivity = $(af^m + bf^2 + c + dt) / [10(1 + \epsilon p)]$ Siemens/meter

t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ϵ = CPcor;

Residual = (instrument conductivity - bath conductivity) using g, h, i, j coefficients

Date, Slope Correction

