

## Pressure Test Certificate

Test Date: **2019-12-18**Description: **SBE-4 Conductivity Sensor**

### Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SBE-4**Serial Number: **2039**

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      MinutesHigh Pressure Test: **8000**      PSI      Held For: **30**      MinutesPassed Test: **True**Tested By: **SW**

High pressure is generally equal to the maximum depth rating of the instrument

