



Sea-Bird GmbH
Postfach 1167, 87401 Kempten, Germany
Phone: +49 831 9 60994 701 Fax: +49 831 960994 709
Email: seabird.eu@seabird.com

Pressure Test Certificate

Customer CNRS/LOV
Job Number E00145
Date 4/10/2013
Technician SW

Serial Number 043678

Low Pressure (PSI) 45 PSI
Time (Minutes) 15 Minutes
High Pressure (PSI) 8000 PSI
Time (Minutes) 30 Minutes

Pass
Fail

Comments

Replaced the main piston "O"-Rings.
Replaced conductivity cell.

