

Pressure Test Certificate

Test Date: **2020-01-14**Description: **SBE-4 Conductivity Sensor**

Sensor Information:

Model Number: **SBE-4**Serial Number: **3678**Replaced the main piston "O"-Rings.
Replaced 3-pin connector.
Replaced main housing.

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** MinutesHigh Pressure Test: **8000** PSI Held For: **30** MinutesPassed Test: **True**Tested By: **PB**

High pressure is generally equal to the maximum depth rating of the instrument

