



Sea-Bird GmbH  
 Postfach 1167  
 87401 Kempten  
 Germany

+49 831 9 60994 701  
 seabird.eu@seabird.com  
 www.seabird.com

SENSOR SERIAL NUMBER: 3598  
 CALIBRATION DATE: 17-Dec-20

SBE 4 CONDUCTIVITY CALIBRATION DATA  
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00120165e+001  
 h = 1.54521276e+000  
 i = -2.45728926e-003  
 j = 2.79304007e-004

CPcor = -9.5700e-008 (nominal)  
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.54913	0.00000	0.00000
-1.0000	34.8757	2.80891	4.97419	2.80892	0.00001
1.0000	34.8727	2.98033	5.08480	2.98033	-0.00000
15.0000	34.8677	4.27725	5.85389	4.27723	-0.00002
18.5000	34.8609	4.62365	6.04264	4.62365	0.00001
29.0000	34.8342	5.70492	6.59676	5.70495	0.00003
32.5000	34.8140	6.07563	6.77615	6.07561	-0.00002

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ε = CPcor;

Conductivity (S/m) = (g + h \* f<sup>2</sup> + i \* f<sup>3</sup> + j \* f<sup>4</sup>) / 10 (1 + δ \* t + ε \* p)

Residual (Siemens/meter) = instrument conductivity - bath conductivity

