

## Pressure Test Certificate

Test Date: **2020-12-18**Description: **SBE-4 Conductivity Sensor**

### Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: **SBE-4**Serial Number: **3677**

### Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** MinutesHigh Pressure Test: **8000** PSI Held For: **15** MinutesPassed Test: **True**Tested By: **BW**

High pressure is generally equal to the maximum depth rating of the instrument

