



SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

Pressure Test Certificate

Customer EMS/France

Job Number 55354

Date 7/21/2009

Technician TW

<u>Serial Number</u>	052048
-----------------------------	--------

Low Pressure (PSI) 50 PSI

Time (Minutes) 15 Minutes

High Pressure (PSI) 10000 PSI

Time (Minutes) 45 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. Upgraded to wet pluggable connector.

