



# SEA-BIRD ELECTRONICS, INC.

1808 - 136th Place Northeast, Bellevue, Washington 98005 USA

Phone: (425) 643-9866 Fax: (425) 643-9954 www.seabird.com

## Pressure Test Certificate

**Customer** EMS/France

**Job Number** 55354

**Date** 8/13/2009

**Technician** AS

**Serial Number** 09P21145-0499

**Low Pressure (PSI)** 45 PSI

**Time (Minutes)** 15 Minutes

**High Pressure (PSI)** 10000 PSI

**Time (Minutes)** 45 Minutes

Pass

Fail

Comments

Replaced the main piston "O"-Rings. Upgraded connectors to wet pluggable

