



Sea-Bird GmbH  
Postfach 1167  
87401 Kempten  
Germany

+49 831 9 60994 701  
seabird.eu@seabird.com  
www.seabird.com

## Pressure Test Certificate

Test Date: **2020-02-06**

Description: **SBE-9plus CTD**

### Sensor Information:

Model Number: **SBE-9P**

Serial Number: **0499**

Replaced the main piston "O"-Rings.  
Replaced main housing.  
Replaced end cap.

### Pressure Test Protocol:

Low Pressure Test: **40**      PSI      Held For: **15**      Minutes

High Pressure Test: **8000**      PSI      Held For: **15**      Minutes

Passed Test: **True**

Tested By: **TK**

**High pressure is generally equal to the maximum depth rating of the instrument**

