



Sea-Bird GmbH
 Postfach 1167
 87401 Kempten
 Germany

+49 831 9 60994 701
 seabird.eu@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 2039
 CALIBRATION DATE: 10-Jan-23

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.03766054e+001
 h = 1.27227149e+000
 i = -3.82888917e-004
 j = 8.71703780e-005

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.85630	0.00000	0.00000
-1.0000	34.7868	2.80242	5.49276	2.80244	0.00001
1.0000	34.7878	2.97376	5.61380	2.97374	-0.00002
15.0000	34.7887	4.26859	6.45528	4.26861	0.00003
18.5000	34.7892	4.61516	6.66235	4.61514	-0.00002
29.0001	34.7872	5.69810	7.27126	5.69810	-0.00000
32.5000	34.7786	6.07016	7.46890	6.07016	0.00000

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

