



Sea-Bird GmbH
 Postfach 1167
 87401 Kempten
 Germany

+49 831 9 60994 701
 seabird.eu@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 3598
 CALIBRATION DATE: 10-Jan-23

SBE 4 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.00172254e+001
 h = 1.54755531e+000
 i = -3.23140272e-003
 j = 3.43569254e-004

CPcor = -9.5700e-008 (nominal)
 CTcor = 3.2500e-006 (nominal)

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (kHz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
0.0000	0.0000	0.00000	2.54915	0.00000	0.00000
-1.0000	34.7868	2.80242	4.97016	2.80243	0.00001
1.0000	34.7878	2.97376	5.08081	2.97374	-0.00002
15.0000	34.7887	4.26859	5.84938	4.26862	0.00003
18.5000	34.7892	4.61516	6.03833	4.61514	-0.00003
29.0001	34.7872	5.69810	6.59358	5.69809	-0.00001
32.5000	34.7786	6.07016	6.77369	6.07017	0.00001

f = Instrument Output (kHz)

t = temperature (°C); p = pressure (decibars); δ = CTcor; ϵ = CPcor;

Conductivity (S/m) = $(g + h * f^2 + i * f^3 + j * f^4) / 10 (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

