



Sea-Bird GmbH
 Postfach 1167
 87401 Kempten
 Germany

+49 831 9 60994 701
 seabird.eu@seabird.com
 www.seabird.com

SENSOR SERIAL NUMBER: 0269
 CALIBRATION DATE: 18-Jan-24

SBE 45 CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -9.783404e-001 CPcor = -9.5700e-008
 h = 1.341702e-001 CTcor = 3.2500e-006
 i = -1.692239e-004 WBOTC = 6.7862e-007
 j = 3.307230e-005

BATH TEMP (° C)	BATH SAL (PSU)	BATH COND (S/m)	INSTRUMENT OUTPUT (Hz)	INSTRUMENT COND (S/m)	RESIDUAL (S/m)
22.0000	0.0000	0.00000	2702.49	0.00000	0.00000
1.0001	34.8534	2.97884	5429.65	2.97879	-0.00005
4.5000	34.8335	3.28619	5635.78	3.28621	0.00003
15.0000	34.7913	4.26887	6248.38	4.26894	0.00007
18.5000	34.7820	4.61431	6449.57	4.61435	0.00004
24.0000	34.7720	5.17276	6761.78	5.17266	-0.00011
29.0000	34.7665	5.69508	7040.99	5.69503	-0.00005
32.5000	34.7624	6.06765	7233.39	6.06772	0.00007

$f = \text{Instrument Output(Hz)} * \text{sqrt}(1.0 + \text{WBOTC} * t) / 1000.0$

t = temperature (°C); p = pressure (decibars); $\delta = \text{CTcor}$; $\epsilon = \text{CPcor}$;

$\text{Conductivity (S/m)} = (g + h * f^2 + i * f^3 + j * f^4) / (1 + \delta * t + \epsilon * p)$

Residual (Siemens/meter) = instrument conductivity - bath conductivity

