

Pressure Test Certificate

Test Date: **2024-02-08**Description: **SBE-45 Mini Thermosalinograph**

Sensor Information:

Replaced the main piston "O"-Rings.
Replaced conductivity cell.Model Number: **SBE-45**Serial Number: **0269**

Pressure Test Protocol:

Low Pressure Test: **40** PSI Held For: **15** MinutesHigh Pressure Test: **40** PSI Held For: **15** MinutesPassed Test: **True**Tested By: **BW**

High pressure is generally equal to the maximum depth rating of the instrument

