

# CLASS PROGRAM

NAME \_\_\_\_\_ ADDRESS \_\_\_\_\_

SCHOOL \_\_\_\_\_ CLASS \_\_\_\_\_

		PERIOD 1	PERIOD 2	PERIOD 3	PERIOD 4	PERIOD 5	PERIOD 6	PERIOD 7	PERIOD 8
TIME	FROM								
	TO...								
MONDAY	SUBJECT								
	ROOM								
	INSTRUCTOR								
TUESDAY	SUBJECT								
	ROOM								
	INSTRUCTOR								
WEDNESDAY	SUBJECT								
	ROOM								
	INSTRUCTOR								
THURSDAY	SUBJECT								
	ROOM								
	INSTRUCTOR								
FRIDAY	SUBJECT								
	ROOM								
	INSTRUCTOR								
SATURDAY	SUBJECT								
	ROOM								
	INSTRUCTOR								

9<sup>3</sup>/<sub>4</sub> in. x 7<sup>1</sup>/<sub>2</sub> in.  
MADE IN U.S.A.

## DRY LAB

\*photocopy and scan at the end of every day  
→ optical engineer notes folder

5/9/13 - make an inventory once full supplies arrive

### Accuri

- filter and tubes change at the end of each leg.

### TODO:

- run all calibrations + create a template
- re-test the thresholds w/ cultures

5/10/13 • changed fluidic bottle filters on:

- sheath fluid
- decontamination fluid
- cleaning fluid

- changed sheath filter (inside instrument)
  - changed pump tubing on both waste (red) and sheath (blue) sides (inside instrument)
- purge air after above changes (pg 28 inst. manual)

- Before cruise starts:
- ✓ replace tubing for pump and filter
  - re-calibrate Accuri (new beads)
  - cal. FlowCAM
  - get all cleaning solutions mixed + in place
- PRINT: - Revised protocols  
- Daily c-lists + routine



DRY LAB

5/11 For FlowCAM:

- S = bead size calibration
- F = fluorescence calibration (says "for Accuri")
- E = Appendix E in FlowCAM protocol - purple
- D, once
- every 2-3 weeks, or beginning + end of log
- measure tubing length
- 10% bleach solution - 10 mL into 90 mL H<sub>2</sub>O
- leave closed, w/ bleach sitting, flush in the morning
- have Marc: look at data - is 50 mL a good amount to run?
- fix lid open

Accuri

- clean SIP w/ squirt bottles
- do a set of cals, inc. size before leaving
- also before leaving, make a new 6- and 8-peak stocks for first

also print out log -  
 CDAM scanner:  
 change filter on wall once a week

★ make sure there is an empty bucket in lab at all times ★

Lee France cell 07 611 88986

From Lee:

- doc for FlowCAM -
- revised Accuri protocol ✓
- list of stuff that celine will get ✓
- log sheets
- daily checklists
- send Lee screen shots and results when possible
- \* culture location ✓
- email Bill Kelley re: insurance dates -

5/12/13

FlowCAM

- tubing length:
- above flow cell = 30cm
- below " " = 37cm
- maybe 0.5cm goes onto pipette tip
- context files for:
  - ① manual Pnme
  - ② sample
  - ③ cleaning
- Note - while water is running, cannot do "setup and focus"




08P87 110 FO see  
see

→ context files are done, need to edit:

- volume to be run

Size cal:


- prime with DIW, until level is 2cm above the bottom of the pipette tip:



- load a run sample context file  
 - change the sample volume to 1 mL  
 - change the basic size acquisition filter (setup > context > filter (use ESB)) to min = 10, max = 2000

- pipette in 1 mL of the 4x size mix  
 - run on auto image mode, save as; cal-size-mix-date  
 - load filters file  
 - make sure the histogram plot is correct.

filters should be: We only have 50! (µm beads)



n

5 95  
10 190

Bottles to label:

ALFA cleaning (amounts are enough for one cleaning)

- ① 1L DIW
  - ② Detergent 50:1 (250 mL<sup>DIW</sup> and 5 mL detergent)
  - ③ Bleach 50:1 (250 mL DIW + 5 mL bleach)
- (DON'T USE STRONGER SOLUTIONS)

CDOM mapper

- |         |                           |                        |
|---------|---------------------------|------------------------|
| ALFA    | ① DIW                     |                        |
| CDOM M. | ② Detergent (5% solution) | 10 mL det., 190 mL DIW |
| FlowCAM | ③ Methanol (50% solution) |                        |
| Accuri  | ④ HCl (10% solution)      |                        |

Ultra FlowCAM  
IFCB sample bottles

- vol? {
- ① Filtered seawater (0.45 µm filter)
  - ② surface Niskin 2 x 250 mL
  - ③ DCM Niskin 2 x 250 mL
  - ④ Bongo surface 50 mL
  - ⑤ Bongo DCM 50 mL
  - ⑥ Bleach 50:1 (0.5 mL bleach + 25 mL DIW)
  - ⑦ DIW

Accuri

- See log sheets
- ① surface Niskin
  - ② DCM

(Flowtime may be taken @ other times)  
 ③ Make 0.2 filtered seawater

→ Daily cal: 8+6 peak beads, sure count + yellow beads  
 Every other day: size cal w/ 4 diff sizes



5/17/13

Made new bead solutions for Accuri:

- 6 + 8 peak
- surecount + yellow standard
- sizes (1, 2, 4, + 10um)

Running all cals, editing protocol (small details)

TO RUN:

• New DIW - 2 min fast

SAVE

• 8-peak

• DIW - 2 min

SAVE

• 6-peak

• DIW - 3 min

SAVE

• count slow

SAVE

• count fast

• DIW - 2 min

• Backflush/unclog X 2

• DIW - 2 min

SAVE

• 1um size (fast and slow)

• DIW

SAVE

• 2um size (fast AND slow)

• DIW

SAVE

• 4um size (fast and slow)

• DIW

SAVE

• 10um size (fast and slow)

• DIW

• Backflush/unclog X 2

• DIW

Cal files: 2 new cells

Count file:

Size file: 8 new cells (2 speeds ea. size)

Bottles to label:

ALFA cleaning (amounts are enough for one)

① DIW

② Detergent 20:1 (250ml DIW + 12.5ml detergent)

③ Bleach 20:1 (250ml DIW + 12.5ml bleach)

(DON'T USE SPONGER SOLUTIONS)

(COM Mixture)

① DIW

② Detergent (20:1 solution) 10ml DIW + 0.5ml detergent

③ Methanol (50:1 solution)

④ HCl (10:1 solution)

⑤ DIW

⑥ Filtered seawater (0.2um filter)

⑦ Surface mistin

⑧ DCM mistin

⑨ Bordo surface

⑩ Bordo DCM

⑪ Bleach 20:1 (0.5ml bleach + 25ml DIW)

⑫ DIW

ALFA

① Surface mistin

② DCM

③ Make 0.5 filtered seawater

④ Amounts are taken @ other times

⑤ Make 0.5 filtered seawater

⑥ Make 0.5 filtered seawater

⑦ Make 0.5 filtered seawater



## TEST RUN OF ULTRAPATH

- 1/7/13 GOMaine water
- ran 1x cleaning cycle
  - reference water (~20psu)
  - \* file name: "test/orient\_r.txt" (r for ref)
  - reference run:
    - int. time up to 385.0 to get a I<sub>MAX</sub> of 75%
    - 200 cm pathlength
    - I<sub>MAX</sub> = 76%
  - ref scan does not look too good... need to use hplc water instead of DLW?

estelledIMIER

- 1/8/13 Bottles still to be labeled:
- For FlowCAM: • FSW (between samples + flushing @ beginning)
- Sample bottles x 6
  - Bleach solution

Accuri: Sample bottles

- 5/19/13 Departure Day
- mixed new cleaning, decon, + sheath fluid for Accuri
  - add 10 cm to FlowCAM input tube length → 0.078 ID
  - had to replace flow cell, was dirty on the outside.
  - \* NOTE - tighten ~~the~~ the screw facing directly at the user on the flowcell holder so that the flowcell will not move.
  - 36 cm below flow cell
  - 30 cm above
  - \* brand new flow cell is quite dirty on the outside of the glass...
  - running bleach to clean inside
  - lens paper used to clean outside of cell
  - Cleaning SIP on Accuri
  - ran ~~6~~ 6 + 8 peak E size each (made new size bead stocks)

At sea:

- empty waste bottle
- empty Accuri waste

\* ASK Celine:  
- disposal of glass tubes? (at sea)



5/24/13 Station #155

LOCAL TIME ~~7:30~~

0730 Turned on FlowCAM P+L, Accuri, Ultracath

0740 Running sheath on Accuri to clear any bubbles. (3x5 min)

Made FSW w/ flo-thru H<sub>2</sub>O

Put fresh HPLC-grade H<sub>2</sub>O in ultracath bottle.

Checked u-path ref salinity: 34 psu

0800 → Local salinity = 35.3

• flow meter for IFCB + alfa showed 0.0 l/min; opened the valve under the bench more

0830 Running 6 + 8 peak beads on Accuri

- Put beads back in fridge, also put in u-path ref. H<sub>2</sub>O

\* Error message on CDM mapper: Run-time error 6:

- emailed Alan, will wait for response. Overflow

0843 Running count standard cal on Accuri

0900 ~~to~~ Rinsed 0.2 μm disk filter w/ 120 mL of hplc water (two 60 mL syringes full).  
wrapped ~~disk~~ filter in foil until use.

0910 Running a cleaning cycle on u-path:

Citranox → MeOH → HCl → hplc H<sub>2</sub>O (wait for bubble; one min each)

0930 Got ref water from fridge; filling wave guide (Ultracath)

0945 Got surface samples for u-path from Rosette  
dist (3x 250 glass amber bot) \*filtered thru  
0.2 μm disk into 3 other bottles



5/24/13 Str 155 cont.  
 1010 Not getting a great ref. scan. Cleaning again.  
 1129 cleaning cycle 4 times. ref. is ~0.0  
 except in the UV - much higher.  
 - running the niskin samples.  
 ask  $\rightarrow$  should the pump be running while the  
 sample scan is going?  
 - COOM spectra #1 looks okay, maybe  
 can just subtract the ref in post-  
 processing?  
 Running:  
 Cleaning  
 Ref \* all samples run  
 Sample 1 w/ 200 cm pathlength  
 Cleaning  
 Ref  
 Sample 2  
 Cleaning  
 Ref  
 Sample 3  
 end: Ref  
 Cleaning  
 - left w/ water in tubing  
 1215 finished @ 1215

cont.  
 1237 Filtered ~~COOM~~ samples thru 300  $\mu$ m  
 mesh w/ FlowCAM  
 plastic funnel.  
 - 4' bottles. 2x 250 mL surface niskin  
 2x 250 mL DCM niskin

- new pipette tip on Flowcam funnel.  
 1241 \* Make filtered SW (from flo-thru) w/ 45  $\mu$ m  
 disk filter + syringe, to flush FlowCAM w/.

1421 Running fcam surf trigger mode  
 1608 Finished running Accuri surf. + DCM  
 (w/ beads in each for conc.)  
 - not too much in water.

FCAM is running, ~30 mins per sample  
 (for 50 mL of sample)

1625 Accuri shut off (finished its automatic cleaning)  
 FlowCAM running. Still to do on it:  
 - DCM auto mode  
 - conc. 0.2 water (net) A mode  
 F mode  
 - conc. 20 water (net) "

\* movement of the boat causes slight  
 shifts in the flowcell + fluid inside...  
 the background calibration is very sensitive,  
 so this means the software starts seeing  
 dirt on flowcell as cells to image. Problematic...



cont.

17:00 email from Alan - he also gets the same error messages; working on a software fix. Also recommends not using Port 4 since it's not pumping + thus could be drawing air into the system.

17:30 Running FCAM conc. 0.2  $\mu$ m water (surf). Most stuff ~~looks~~ is small, rarely the chains we saw w/ the surf. riskn. maybe cells are damaged in the conc. water process?

18:00 Finished 0.2 conc. water. Did not filter thru 300  $\mu$ m net. Started 20  $\rightarrow$  water (net). Did filter thru 300  $\mu$ m net. - much more big stuff / chains in 20  $\rightarrow$  water.

18:35 Finished 20  $\rightarrow$  net water. ~~Many~~ Many phyto. cells (+ chains)

18:54 Trying to get IFCB into good focus. (using the focus motor under "hardware" tab.

19:31 Running meso (~800 m depth) samples in FlowCAM... trigger mode = basically zero cells (did not filter first meso bottle... did filter second "A" bottle... do not expect a diff. for this reason.)

cont.

20:27 running meso on auto mode

- test on CDOM mapper for ALAN: run sample in place of ~~flow~~ ref H<sub>2</sub>O.

LOCAL TIME Station day 2

0851 CDOM mapper crashing frequently. Screen shots sent to Alan.

0900 Started a new ALFA file: "alfa-20130525-070000"  
 $\rightarrow$  made on 5/17  
Running peak + count calcs on Accuri.  
- Also size calcs (solutions made ~~on~~ on 5/19

Fluor. cal on FlowCAM...

0921 CDOM mapper crashed again. (all new databases created since cruise start on 5/19 are due to software crashes)

0927 looks like some bubbles in IFCB images...  
- stopped ~~acc~~, running "debubble and refill"  
- re-started

0935

Accuri 6+8 peak runs look ok, not perfect. Try making new bead solutions tomorrow.

Running Accuri ~~size~~ col count - looks good



1014 Running accuri size cal. looks not quite as good as prev. runs - make new size stocks for next size cal.

1046 size 4µm beads are dramatically reduced in count.  
→ discarding all bead solutions.

1126 cleaning Ultraphath  
- took "dark scan" (w/shutter off)  
and "reference scan"

5/26/13 Ran Ultraphath w/ flow-thru 1100 local time

5/27 Ran Ultraphath w/ flow-thru 1800 local time

5/28 1200 - made new reference water, replaced old ref water in CDM mapper, also added new HPLC water. Started a new CDM mapper database.

- Crashed again... before crashing, spectra did not look good.

14:51 took a flow-thru sample  
- made new 6+8 peak stocks,  
size stocks, daily count standards,  
- Running Accuri cal.

local time

1800

Ran Ultraphath, w/ cont. flow water  
temp = 10.39°C

sal = 35.3

sal ref = 35 psu

200 cm pathlength

77% para

385 int. time.

lat = 59.379°N

lon = -9.3258

Accuri:

- ran all cal: peaks, counts, sizes  
- doesn't look like too much in the water

sample from cont. flow

2min FL3 slow: 26.3 µm

" " fast: 130.8 µm

" FSC slow: 26.3 µm

Flow cam: running 200 mL @ 1.5 mL/min  
time start: 14:20 (from flow-thru)

duration: 2 hrs 12 min

# of particles: ~~28,770~~ 28,770

vol <sup>processed</sup> ~~sampled~~: 199.6

ESD: mean = 27.25

min = 10.00

max = 99.94

fluid vol imaged = 29.644 mL

1.5 mL/min



To DO on wed 5/29:

- 50 mL @ diff. speeds in FlowCAM test.
- Run Accuri - first make sheath fluid
- Run ~~extra~~ ultrapath (new ref water + HPLC water for cleaning)
- FlowCAM fluor. cal?
- Accuri cryotals?
- make ALFA cleaning solutions
- test of new vs. old ref water \*

5/29/13

Took flow thru water:

0802 UTC

60.6374°N

-8.2697°W

t = 9.8233°C

S = 35.3 psu

~~0830 started~~

UTC time from now on

0830 - started a 50 mL @ 1.5 mL  
run w/ flow thru water on mm-1  
FlowCAM (speed test)

0907 started next 50 mL run on  
FlowCAM, @ 0.75 mL/min.

- made new Liter of Accuri  
sheath fluid

- made bleach + detergent solutions  
for cleaning ALFA.

5/29

sent ~~ALFA~~ ALFA screen shots to Mark,  
and accuri screen shots to Lee

1020 - Started zooprocess for yesterday's  
200 mL run and today's 2 x 50 mL  
runs.

1217 - mixed new ref water w/ fresh HPLC  
water; 37 psu local sol = 35.3 psu

1253 - running ~~count~~ peak ~~count~~  
cals on Accuri  
- turning on + cleaning U-path.

1324 - running count cals on Accuri

- FlowCAM, upath, Accuri ready for resette  
H2O

1407 started running FlowCAM Surface  
Auto-mode, 200 mL, 1.5 mL min<sup>-1</sup>

1426 - running ultrapath  
IMAX = 75%

1443 - running Accuri samples

1503 - finished running ultrapath + Accuri



5/30 Faroe Islands

0700 Cleaned ALFA (flow-thru is stopped)

5/31/13 Underway again from Faroe Is.

~~1700~~ 1700 Re-started flow-thru + all instruments, except CDom mapper. Still having problems, in touch w/ Alan Hails.

FOCUS MOTOR:

2012 → 350

6/1/13

1000

Email from Alan w/ several things to try on CDom mapper... no luck; still having error message/software crashing. Also, now the software seems to stop and we get "not responding" in task manager at the first reference cycle.

1010

Starting a new ALFA file w/new settings:  
PDP shots = 75  
Blue Int. = 0.7  
Green Int. = 3

"alfa\_20130601-101000"

No sample today - waiting for int'l w

Next to do:

Accuri

- 5 min sample runs
- extended cleaning cycle

FlowCAM

- try 100 mL (max raw images = 99999)
- process + sort data ✓
- Fluorescence cal

- inventory of glass vials
- send Lee IFCB images

1500

Power outages on all instruments.  
Re-started ALFA  
Re-started IFCB

0800

STATION 157

- mixed new ultrapath water
- local sal = 35.0
- ref sal = 35 psu

\* PD

Accuri ex. cleaning @ end of the day.

- started Accuri

- changed all re-useable tips

0820

- started a new ALFA file



6/2/13

- replaced white pad under Accuri SIP

0900 Running FlowCAM fluorescence cal

0920 Running Accuri calcs

- very slight doubling in peak calcs  
- count calcs look good

1000 FlowCAM fluor. cal. stopped at 30 mins.,  
the context file overrides the  
stop conditions (apparently)

1005 Started running FlowCAM Auto-image.  
100 mL  
1.5 mL/min  
99999 max raw images

1045 starting Accuri samples

11:20 Finished Accuri  
started FCAM fluorescence run  
- lots of tintinids

1142 Running extended clean of  
flow cell on the Accuri  
~~did not hear pumps running,  
and then time~~

6/2/13

1314

Accuri restarted w/ DLW  
(message for extra start-up time = expected)

- Ultrath path start running  
- Processing FlowCAM images w/ Zooscan  
process

1326 Accuri shutdown

W \* Added the CDM barcode to the comment  
line on the TIDAS DAQ software.

1410 Finished ultrath path

\* changed 0.2  $\mu$ m filter on wall for CDM  
mapper

CDM mapper cleaning:  
H<sub>2</sub>O, det, air, meth, HCl

1820

CDM mapper:

Ran for ~20 min to get bubbles out  
of new 0.2  $\mu$ m filter. took off orange  
caps; tubes just in bottles w/ parafilm  
on top. Primed all ports; started  
\* Run-time error 13 at end of cleaning  
sequence.  
cycle (before ref)  
- changed all cleaning to "0" in parameters  
(and cycles between cleaning = 100).  
- manually collecting dark + ref, then set  
of sample spectra.







6/4/13  
1115

Finished U-path w/ flow thru.  
\* Every day from now on (except fill station or CTD cast):  
Take a flow thru sample @ the same time as HPLC, DOC, and nutrients (so, those three plus Accuri, FCAM, + U-path are all run w/ same thro).

1348 Zooprocess validation from 6/2/13

Raw images:

# of Anknids: 8, 8, 8, 2, 5, 6, 6, 7, 10, 9, 10, 5, 9 93 tot.

File = flowcam\_20130602-stn157-S\_A (raw)

Work images:

# of Anknids: ### ### ### ### ### ### ###

### ### ### ### ### ### ### 70 tot

file = 20130602-stn157-s-a

Meso FlowCAM:

67 10.05 6905

0 13.064

B02 = ~~400~~ vol conc

45.4 vol pump

comment = ~~40000~~

D02

50 vol,

B20 = ~~50~~ vol. conc.

xx vol. pump

~~70000~~ comment

ll

6/5/13 Station 159 (short)  
= surface only

1330 Made ref water

- started CDM mapper

1345 → crashed, error to overflow  
- restarted

1350 calibrating Accuri (peaks)

1600 count cal + sample (Accuri)

1615 FCAM Auto started

CDM mapper is still going; spectra look generally correct but noisy (?)

1650 Did de-bubble + refill on IFCB; restarted CDM

1720 After de-bubble + refill, positions in the ROI vs. Peak plot look good.

FCAM capture x vs capture y looks good. Also, changed the min. diameter to 20 μm... 10 gets too much junk (unidentifiable)

1857 CDM mapper still running, data looks okay? proper general shape but noisy?



\* check flowcam & comp. free space!!

Time: ~~clean ALFA~~ (when flow thru is stopped)

- send screen shot to LKB
- validated predicted vignettes
- email re: com mapper

6/6/13 Station 160 (shard station)

0700 made new ref water (and replaced  
psu = 35 CDM mapper  
local = 35.17 ref + HPLC-H<sub>2</sub>O)

0725 Prepped bottles w/ stickers, got new empty log sheets  
Turned on Accuri

0730 Running Accuri peak calcs  
U-path ref water dishes

0745 checked IFCB images

0755 IFCB focus:

350 >> 700 pos  
350 >> 1050 pos  
350 >> 1400 pos  
350 >> 1750 pos

(all objects now out of focus, going back the other way)

Back to zero pos →

IFCB focus:

6/6/13

- 350 pos
- 700 pos
- 1050 pos
- 1400 pos
- 1750 pos

0800 Rosette arrive - go to get H<sub>2</sub>O + filter CDM

0825

- 2100 pos
- all objects out of focus
- 1400
- 1050 some in focus now
- 700
- 350
- 300
- 250
- 200 getting better
- 150
- 100



0830 Turn off v-path, start cleaning cycle. First put Accuri + FCAM samples in fridge.  
Added new MeOH + citranox to cleaning bottles (pour outside)

0841 IFCB - made small motr ~~sp~~ step = 20

- 80 pos
- 60 pos Some still out of focus
- 40
- 20 most in focus
- 0 pos



6/6/13  
0853

IFCB focus (u-path running also)  
+ 20 pos  
40 pos  
Small step 10  
50 pos most things in focus (best yet)  
60 pos  
50 pos  
Small step size = 5  
45 pos  
40  
35

0930 Finished u-path. Dumping waste.

0940 Started FCAM-A (made filt. seawater for primary + filtered sample thru 300  $\mu$ m mesh).  
Making Accuri count standard  
- filtering accuri sample + FSW

0950 Started a new ALFA file  
Running count + size calcs on Accuri

0955 IFCB focus end on pos = 55  
- check processed image files from after this time later.

1100 Running FSW + Samples on Accuri.  
Running FCAM Trigger mode  
Running CDAM mapper w/new .exe!  
- spectra look pretty good, emailed Alan one.

6/6/13

1300 After lunch - did all lab dishes + started Zooprocess on FCAM computer.  
\*Always check disk space on FCAM comp. before running samples\*

→ new Z: drive on FCAM computer - ~~for~~ this is the back-up drive. In general, work on the C: drive, which is the local FCAM comp.

158 0603 - 0746 67.082  
W 0°.150

A-20 SRF 37.04 mL

6/7/13 Same as yesterday. - short station 161  
Made new sheath fluid for Accuri.  
Did a longer manual cleaning on CDAM mapper.

0944 IFCB debubble + refill



6/8/13 ~~Star~~ Short station 162

0820 CDM mapper - water <sup>(fluids)</sup> appears to be going in but not coming out... open later to look inside.

0930 - made new Accu size stocks  
- peak cals have doubling, tried backflush + unclog, did not work, removed SIP to clean top w/ squirt bottle.  
Running DIW for 15 min (did 2x B-flush / unclog first)

- Accu peak cal looks a little better after SIP removal + cleaning

6/9/13 Long Station #163

0630 made ref water - 35 psu  
local = 34.8  
temp = 2.1°C

0700 New ALFA file  
Checked IFCB images - lots of blobs, mostly small single cells, many not in focus.

CDM mapper: opened yesterday, is leaking inside somewhere. Unplugged per Alan's instruction + waiting on him for more info.

6/9/13

0922 "Backflush with sample" on IFCB (stop acq. first)

0940 Checked IFCB camera - screen was black, \* stopped acq., unchecked backflush, start again.  
\* nevermind - checked "view images" and it was fine.  
LOTS of blobs post-backflush!

Many particles in the water today - for the first time I had to use two filters to get thru the CDM filtering

1612

Running FCHM. File naming error:  
flawcam-20130609-stn163-S-A SHOULD BE:  
flawcam-20130609-stn163-B20-A  
AND: flawcam-20130609-stn163-S-A-true is the real file of that name.

1730

Make new 6+8 peak beads.  
Flipped top O-ring on SIP.  
Re-running cals.  
Looks bad - triplets.  
Tightened SIP collar. Looks good!  
Re-running FSW + sample.



6/10/13 2<sup>nd</sup> day of long station - only meso water from rosette.  
Using cont. flow water for Accuri, U-path, FCAM (also meso in FCAM)

- Made new Accuri sheath fluid

Lab room is hot today. (heaters are on)

Accuri calcs still look good.

U-path ending IMAX was 83%?

Why so much higher? Ask Jeanne.

6/11/13 Daily cast will be @ ~2 pm

0923 Running extended cleaning solution on Accuri

1600 Noticed IFCB was not restarted after last night's AC-S cleaning - start aeg.

U-path reference spectra look good.

AC-S filtered flow is back up to ~7 l/min (was down to ~1 yesterday) after new filter.

6/12/13 Short station. #165

0730 Made new ref. water.

0800 Started up Accuri, running peak calcs. Added more decon + cleaning fluid to bottles (did not mix more; already had some in mixing bottles)

- mixed more diluted bleach for FCAM cleaning.

0900 U-path ref is exactly at zero all across...?

→ axis values were off - needed click on the small "LY" button to make it pink instead of the "LX" button being pink

- had to click the button again. Data + ref. look nice.

1030 ~~data~~ IFCB debubble + refill; backflush w/sample ~~for zone syringe~~

6/13/13 LAST station! : #166 (Daily/short)

w/backflush left on on IFCB, data look good... maybe this will help?





- 6/13/13 - List of things to tell next dry lab person → Dry lab report
- FlowCAM # of vignettes

6/15/13 w/ Joannie in Tromso

- 1) - ALFA cleaning &
- 2) - COOM mapper (get it working?)
- 3) - FlowCAM (change flowcell + calibrate) <sup>+ syringe cleaning</sup>
- 4) - Accuri - change filters + pump tubing
- 5) - Ultracath data
- 6) - chemical inventory + dilutions

DAIFA: 2 cleaning cycles. The 2<sup>nd</sup> was done to try to get a flatter spectra as we still could see a little peak ~ 625nm

the 2<sup>nd</sup> cleaning cycle improved it a bit but still see that little peak

Mark from Ali's email & Alexander think it could be due to air in the instrument. They suggested filling the instrument with milliQ water via gravity (no pump) & milliQ bottle as high as possible. Joannie & I will try it today (17 June)

2) COOM Mapper: its works → We tried with sampled offshore water (low COOM) and were not successful. We added the <sup>SEC</sup> feci drops and got a beautiful spectra. Then diluted with filtered sea water.

2X The <sup>same</sup> sample (sea water + feci drop) was analyzed with ultracath (50cm path length) & same shape for spectra but relative values were a little higher with COOM Mapper.  $C_{410nm}$ : COOM mapper 1.17  $\cdot 1/m$  ultracath 0.7  $\cdot 1/m$

we noticed with COOM mapper unconstant (no replicable) spectrum within its 10 <sup>samples</sup> cycles

⊖ For unknown reason, do not continuously work? software

3) FlowCAM: <sup>we</sup> change the flowcell & clean the syringe & filled the bleach 501<sup>m</sup> bottle & calibrate for size & count beads. - created a new context file: run\_sample\_context.tromso.ctx

4) Accuri: <sup>we</sup> changed filters into bottles + inside Accuri. we changed tubing into 2 pumps inside Accuri. we will (17 June) look at results and Ali will demo.

5) Ultracath data looked good.



6) Chemicals: 25L Citranox to be diluted

5L Methocel or more to be diluted

(Dilute) Fill up the 5L 250mL bottles 0.1N (or 1%)



See Mats, Geraldine, Sophie and/or Marianne at NPI  
or Gerald or Claudia at Akvaplan for further help