



# **WETView 7.0**

## **User's Guide**

WET Labs, Inc.  
P.O. Box 518  
Philomath, OR 97370  
Tel: 541-929-5650  
fax: 541-929-5277  
[www.wetlabs.com](http://www.wetlabs.com)





---

## Table of Contents

|  |    |
|--|----|
| 1. Introduction.....                           | 1  |
| 2. System Requirements and Installation .....  | 1  |
| 2.1 Installing the Software .....              | 1  |
| 3. Quick Start.....                            | 1  |
| 4. Operation Overview.....                     | 2  |
| 4.1 Controlling WETView.....                   | 2  |
| 4.2 How Data is Collected .....                | 2  |
| 4.3 Displaying Data .....                      | 3  |
| 4.4 Rescaling Graphs .....                     | 3  |
| 5. Collecting, Graphing, and Saving Data ..... | 4  |
| 5.1 WETView Menus .....                        | 4  |
| 5.2 Opening a Device.....                      | 5  |
| 5.3 Opening and Saving a Data File .....       | 5  |
| 5.4 Collecting Data .....                      | 6  |
| 5.5 Graphing Data.....                         | 6  |
| 6. Configuration .....                         | 8  |
| Appendix A—ac-9 Meter Files.....               | 11 |
| ac-9 Configuration File.....                   | 11 |
| ac-9 WETView Data Files.....                   | 11 |
| ac-9 Raw Data.....                             | 14 |
| Appendix B—ac-s Meter Files .....              | 18 |
| ac-s Configuration File .....                  | 18 |
| ac-s WETView Data Files .....                  | 20 |
| ac-s Raw Data.....                             | 22 |



## 1. Introduction

WETView is a data acquisition and display program for acquiring and viewing the data produced by WET Labs absorption and attenuation instruments. It runs on IBM PC compatible computers. This document describes what is required to run WETView, how to install it and how to use it.

## 2. System Requirements and Installation

It is recommended you have at least a 90 Mhz Pentium with at least 16 Mb of memory. You will need about 3 Mb of disk space for the WETView program and related files. The speed of the processor limits how much real-time graphing can be accomplished without data loss. The size of memory limits the number of data points that can be plotted in one run. WETView requires Windows 95 or newer operating system.

The WETView CD contains five files: `SETUP.EXE`, `Wethsx.001`, `Wethsx.002`, `*.dev` and `*.cal`.

Once installed, the WETView directory contains the following files:

- `setup.exe` WETView installation executable
- `uninst.dll` File associated with uninstall utility
- `uninst.exe` The uninstall utility
- `uninst.lrm` File associated with uninstall utility
- `Wethsx.001` File associated with program file
- `Wethsx.002` File associated with program file
- `WetView.exe` WETView program
- `WETVIEW.UIR` WETView user interface resource file
- `*.DEV` Device configuration files (the actual name(s) of these files depend upon the instrument(s) that you have purchased)

### 2.1 Installing the Software

1. Insert the CD.
2. Open the CD and double-click on the `SETUP.EXE` icon. `SETUP` will guide you through the rest of the installation process.
3. Copy the `*.dev` and `*.cal` files to the hard disk. WETView initially looks for the `.dev` file in the installation directory (default is `c:\Program Files\wetlabs\wetview7`).





### 3. Quick Start

This section gives instructions for you to quickly get a single meter connected and displaying data. After installing the software:

1. Connect the meter to an appropriate power supply.
2. Connect the data line to one of the communication ports on your computer.
3. Turn the meter power supply on.
4. Start the WETView program.
5. Open a device configuration file by pressing the **^O to Open** button at the top of the screen (or by typing control-O). A dialog box will appear, allowing you to select a file to open.
6. Select the COM port to which you attached the meter.
7. Press the **F1 to Start** button at the top of the screen.
8. After several seconds, stop the acquisition by pressing the **F2 to Stop** button.
9. Save the newly acquired data to a file. Name the file **TMP.DAT**.
10. Rescale the graph by selecting the **All channels** button in the lower right corner of the screen. You can zoom in on one channel by press the colored button next to the channel's name on the right-hand side of the screen.
11. Open the data file you have just created by selecting **File/Open .DAT file...** from the File menu. Choose the file named **TMP.DAT**.
12. View the data in the file by pressing the **"F1 to Start"** button at the top of the screen.

## 4. Operation Overview

WETView acquires data from WET Labs' ac meters and displays collected data. WETView can perform these tasks simultaneously or independently.

### 4.1 Controlling WETView

You may use a mouse or keyboard commands to control WETView. The assigned function keys are:

|    |  |
|----|--|
| F1 | Start acquiring data                     |
| F2 | Stop acquiring data                      |
| F3 | Bring up Graph Options dialog            |
| F4 | Bring up Binning Options dialog          |
| F5 | Move data on graph to the left (or down) |
| F6 | Move data on graph to the right (or up)  |
| F7 | Zoom graph in                            |
| F8 | Zoom graph out                           |

^O (control-O) Opens a device for data acquisition.

To operate controls on the screen using the keyboard, select the control by tabbing until the desired control is highlighted, then change its value. The method for changing a control's value depends on the type of control:

- For a text or numeric field, type the desired value.
- For a radio button, type a space to toggle its value.
- For a button, type Return to activate the button.

Most dialog boxes have an OK and a Cancel button. The function key F1 will select OK and the Esc key will cancel the dialog.

### 4.2 How Data is Collected

As data is collected, it is averaged into bins. The average of each bin is stored and can be written to a file after the acquisition is complete. The number of samples per bin is set by the user via the Options/Channels/Binning... menu item. Once the data has been averaged into bins, those bins can be further averaged into bins of bins for plotting.

Regardless of which channels are being displayed, all data acquired from an instrument is stored and can be saved to a data file.



### **4.3 Displaying Data**

Data can be read directly from an instrument and displayed in real time, or it can be read from a previously stored data file. You may configure the X and Y ranges for graphs and select which traces are to be plotted. Collected data may be averaged into bins for display. Thus, you can average acquired data into bins of 4 samples each, and then display bins of 12 samples each by selecting a collection bin size of 4 and a display bin size of 3.

### **4.4 Rescaling Graphs**

When data is displayed in a graph, you can rescale the graph's bounds to zoom in on a single channel by clicking on the small button associated with the channel's label (on the right side of the screen). Clicking on the **All channels** button will rescale the graph so that all displayed channels will be shown.


In the lower right corner of the screen are four buttons for navigating around a displayed graph. You can zoom in (magnify) a graph, or zoom out.


You can also move the bounds on the graph to shift the viewed data to the right or left on the current graph (for strip charts, these move the graph up and down).


You can change other characteristics of a graph by selecting **Options/Graph** (or, equivalently, the **Graph/Options** menu item).

## 5. Collecting, Graphing, and Saving Data

### 5.1 WETView Menus

|   |  |     |   |             |  |                |   |         |   |           |  |      |                            |
|---|--|-----|---|-------------|--|----------------|---|---------|---|-----------|--|------|----------------------------|
|  | <table border="1"> <tr> <td>New</td> <td>Clears the current graph and disconnects from the current instrument.</td> </tr> <tr> <td>Open Device</td> <td>Chooses which type of device to collect data from.</td> </tr> <tr> <td>Open .DAT File</td> <td>Opens a previously saved data file for display.</td> </tr> <tr> <td>Save as</td> <td>Saves the most recently acquired data into a data file.</td> </tr> <tr> <td>Configure</td> <td>Allows you to edit the configuration for an instrument, including the calibration constants.</td> </tr> <tr> <td>Quit</td> <td>Quits the WETView program.</td> </tr> </table> | New | Clears the current graph and disconnects from the current instrument. | Open Device | Chooses which type of device to collect data from. | Open .DAT File | Opens a previously saved data file for display. | Save as | Saves the most recently acquired data into a data file. | Configure | Allows you to edit the configuration for an instrument, including the calibration constants. | Quit | Quits the WETView program. |
| New   | Clears the current graph and disconnects from the current instrument.  |     |   |             |  |                |   |         |   |           |  |      |                            |
| Open Device   | Chooses which type of device to collect data from.   |     |   |             |  |                |   |         |   |           |  |      |                            |
| Open .DAT File  | Opens a previously saved data file for display.  |     |   |             |  |                |   |         |   |           |  |      |                            |
| Save as   | Saves the most recently acquired data into a data file.  |     |   |             |  |                |   |         |   |           |  |      |                            |
| Configure   | Allows you to edit the configuration for an instrument, including the calibration constants.   |     |   |             |  |                |   |         |   |           |  |      |                            |
| Quit  | Quits the WETView program.   |     |   |             |  |                |   |         |   |           |  |      |                            |

|  |  |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
|--|--|----------------------|--|---------------------|---|------------------|---|----------------------------|--|-----------------------------|--|--------------------------|--|---------------|---------------------------------|---------------|------------------------------------|----------------|--|----------------|--|
|  | <table border="1"> <tr> <td>Absorption vs. Depth</td> <td>Selects the vertical graph of absorption versus depth.</td> </tr> <tr> <td>Absorption vs. Time</td> <td>Selects the vertical graph of absorption versus time.</td> </tr> <tr> <td>Time Strip Chart</td> <td>Selects the horizontal strip chart for display (absorption vs. time).</td> </tr> <tr> <td>Absorption vs. Temperature</td> <td>Selects the vertical graph of absorption versus temperature.</td> </tr> <tr> <td>Spectrograph (Scatter plot)</td> <td>Selects the scatter plot of absorption/attenuation vs. wavelength.</td> </tr> <tr> <td>Spectrograph (bar chart)</td> <td>Selects the spectrum display of absorption/attenuation vs. wavelength.</td> </tr> <tr> <td>Open Graph...</td> <td>Opens a graph file for viewing.</td> </tr> <tr> <td>Save Graph...</td> <td>Saves the current graph to a file.</td> </tr> <tr> <td>Print Graph...</td> <td>Prints the current graph to a printer.</td> </tr> <tr> <td>Options...(F3)</td> <td>Allows you to configure the scaling of graphical output.</td> </tr> </table> | Absorption vs. Depth | Selects the vertical graph of absorption versus depth. | Absorption vs. Time | Selects the vertical graph of absorption versus time. | Time Strip Chart | Selects the horizontal strip chart for display (absorption vs. time). | Absorption vs. Temperature | Selects the vertical graph of absorption versus temperature. | Spectrograph (Scatter plot) | Selects the scatter plot of absorption/attenuation vs. wavelength. | Spectrograph (bar chart) | Selects the spectrum display of absorption/attenuation vs. wavelength. | Open Graph... | Opens a graph file for viewing. | Save Graph... | Saves the current graph to a file. | Print Graph... | Prints the current graph to a printer. | Options...(F3) | Allows you to configure the scaling of graphical output. |
| Absorption vs. Depth   | Selects the vertical graph of absorption versus depth.   |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Absorption vs. Time  | Selects the vertical graph of absorption versus time.  |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Time Strip Chart   | Selects the horizontal strip chart for display (absorption vs. time).  |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Absorption vs. Temperature   | Selects the vertical graph of absorption versus temperature.   |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Spectrograph (Scatter plot)  | Selects the scatter plot of absorption/attenuation vs. wavelength.   |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Spectrograph (bar chart)   | Selects the spectrum display of absorption/attenuation vs. wavelength.   |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Open Graph...  | Opens a graph file for viewing.  |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Save Graph...  | Saves the current graph to a file.   |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Print Graph...   | Prints the current graph to a printer.   |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |
| Options...(F3)   | Allows you to configure the scaling of graphical output.   |                      |  |                     |   |                  |   |                            |  |                             |  |                          |  |               |                                 |               |                                    |                |  |                |  |

|   |   |  |    |  |                      |    |  |
|---|---|--|----|--|----------------------|----|--|
|  | <table border="1"> <tr> <td>Graph...</td> <td>F3</td> <td>Allows you to configure the scaling of graphical output.</td> </tr> <tr> <td>Channels/ Binning...</td> <td>F4</td> <td>Allows you to set the bin size for sampling and the traces to display.</td> </tr> </table> | Graph...   | F3 | Allows you to configure the scaling of graphical output. | Channels/ Binning... | F4 | Allows you to set the bin size for sampling and the traces to display. |
| Graph...  | F3  | Allows you to configure the scaling of graphical output.               |    |  |                      |    |  |
| Channels/ Binning...  | F4  | Allows you to set the bin size for sampling and the traces to display. |    |  |                      |    |  |

## 5.2 Opening a Device

Select **File/Open Device...** to open an instrument for data collection. (You are opening a device configuration file that describes the device to be used.)

Once you have selected an instrument, you will be asked to identify the source of the data stream. This can be either a communication port that the device is connected to or a binary file. (A binary file is created by using a communications program to read a device's data from a serial port and write it directly to a file in binary format.) Choose **COM1**, **COM2**, or **.RAW File**.

Once you have selected the appropriate port(s), WETView attempts to synchronize with the device on that port. If there is any error in doing this, it will be reported immediately. Possible errors are:

|                        |   |
|------------------------|---|
| Port timed out         | WETView timed out trying to read data from the selected port. Was the correct port selected? Are cables and connections correct?                                  |
| Bad format from source | WETView did receive data from the selected port, but it did not recognize the data stream. Is the correct device connected to the selected port? Is it turned on? |

Once WETView connects to the instrument, it checks to see if the serial number of the instrument matches the serial number given in the configuration file. If it doesn't match, you are warned, but can elect to continue.

### Caution

If the serial numbers do not match, the calibration information stored in the configuration file probably will not be correct for the instrument, so collected data is suspect.

**! Software Error:** Current version may always display this warning regardless of matching serial numbers.

## 5.3 Opening and Saving a Data File

Select **File/Open .DAT File...** to open a WETView data file for graphing. Choose the file you want to open. Once the file is open, select **Start** from the **Acquire** menu (or type F1). Possible errors in opening the file:

|                 |   |
|-----------------|---|
| File not found  | The file you named does not exist. Did you type the name correctly?                                       |
| Bad file format | WETView did not recognize the format of the file. Did you select a data file that was created by WETView? |



|                            |  |
|----------------------------|--|
| Strange error opening file | WETView ran into an unanticipated error. Did you select a data file that was created by WETView? |
|----------------------------|--|

Select **File/Save as...** to save the most recently acquired data to a text file. By convention, data files have the extension **.DAT**, though this is not required.

As WETView collects data, it writes that data to a temporary file called **WETVIEW.TMP**. When you request to save the data, the file is renamed to the name you specify. If you quit WETView without saving the data, the **WETVIEW.TMP** file is not deleted, so you can manually rename it if you wish.

#### **Note**

Because the temporary file, **WETVIEW.TMP**, is renamed when you save it, the file cannot be saved to a disk other than the one that contains the current directory.

---

## **5.4 Collecting Data**

Press the **F1 to Start** button to begin acquiring data from the currently opened device. This item is highlighted only if you have a device open. The function key **F1** works as well. As data is acquired, the traces you have selected (via the **Options/Channels/Binning** menu) are graphed. Graphing the data points uses memory and for long collection runs WETView may exhaust all available memory (this is not true for the Strip Chart graph). If WETView runs out of memory, it erases the graph and begins graphing from where it left off. All collected data is saved in a file, whether or not it was erased from the graph.

### **5.4.1 Changing the Data Collection Options**

Select **Options/Channels/Binning** to select which traces to display and to set the collection and display bin sizes.

### **5.4.2 Ending Data Collection, Exiting WETView**

Once acquisition begins, an **F2 to Stop** button is displayed near the top of the screen. Selecting that button with the mouse, or pressing the **F2** key stops data acquisition. Select **Quit** to quit the WETView program.

## **5.5 Graphing Data**

Select **Graph/Open Graph...** to open a previously saved graph file. By convention, WETView expects graph files to have a **.GPH** extension.

Select **Graph/Save Graph....** to save the current graph to a file. The graph may be opened later and viewed, rescaled, and printed.

Select **Graph/Print Graph....** to print the currently displayed graph.

### **5.5.1 Selecting the Graph to Display**

WETView provides four types of graphs:

1. A vertical X-Y graph of absorption versus depth
2. A vertical X-Y graph of absorption versus time
3. A horizontal strip chart of absorption versus time
4. A vertical graph of absorption versus temperature.

There are trade-offs to be considered when choosing which display to use.

The X-Y graphs accurately plot either time and absorption or depth and absorption, whereas the strip chart accurately plots absorption, but uses a constant time base. However, the strip chart scrolls continuously as data is collected, so you can have detailed view of the most recently received data, whereas the X-Y graph is static. When the data runs off the graph, it is not visible until you stop data acquisition and rescale the graph. In either case, all data acquired is logged to the data file.

The absorption versus temperature graph is used for temperature calibration. The temperature reported is the temperature internal to the instrument, not the ambient temperature. Also, since the temperature is reported once for each set of ten data lines, all ten data lines will be plotted at the same temperature. We recommend setting the collection bin size to a multiple of 10.

### **5.5.2 Changing Graph Options**

Select **Graph/Options...** to change the appearance of the currently displayed graph. For X-Y graphs, you may set the X and Y ranges and the number of divisions for each axis. For strip charts, you may set the Y range and the speed that the chart scrolls.

## 6. Configuration

The ac-9 measures the absorption and attenuation of 9 different wavelengths of light; the ac-s measures from 80–90. Configuration is specific to each type of device. Each of the measured quantities has a calibration offset added to it so that the resulting number is the difference between a clean water measurement and the measurement in the field. You can manually set the value of this offset for each channel; however, determining the correct value to use is not trivial.

### Caution

We recommend you change **only** the instrument description and the color used for plotting it. Changing the baud rate will most likely cause WETView to be unable to communicate with the instrument. Changing the calibration constants will affect the accuracy of your results.

---

Usually, a “clean water calibration run” is performed. Chemically pure water is run through the instrument and the values for absorption and attenuation are averaged over a number of samples. These values are then used as the calibration offsets.

You can change the following settings for the ac-9 meter:

- calibration offsets for absorption measurements
- calibration offset for depth measurement

You can change the following settings for the ac-s meter:

- calibration offsets for absorption/attenuation measurements



To change an instrument’s calibration, a new device file must be collected. The steps of this process are outlined below.

1. Ensure the instrument is properly connected to a power supply and host computer, and that WETView is properly loaded and operational. Open the Configuration screen by selecting **File/Configure...** Choose a device file. The window below (either ac-9 or ac-s) will appear.


**AC - Meter Configuration**

Description: AC9 Absorption and Attenuation Meter

**Trace Information**

Label: a650

Value Offset: 7.74761

Plotting Color: 

**Channel Calibration**

Samples to average: 20

A Channels  C Channels

Buttons: Zero All, Auto-Cal

**Meter Information**

Type: 0 Number: 272

Pathlength: 0.25000

Depth Coeff: 0.01700

Baudrate: 19200

**Depth Calibration**

Depth Offset: -3.19900

Buttons: Auto-Cal, Cancel, OK

ac-9 configuration screen

**AC-Spectra Configuration**


Description: ACS Meter

**Trace Information**

Label: 398.568

C Offset: 1.69175

A Offset: -0.80038

Plotting Color: 

**Channel Calibration**

Samples to average: 40

A Channels  C Channels

Buttons: Zero All, Auto-Cal

**Meter Information**

Type: 53 Number: 2

Pathlength (meters): 0.25000

**Device Settings**

Baudrate: 115200

Buttons: Set Now

**Depth Calibration**

Depth Offset: 0.00000

Buttons: Auto-Cal, Cancel, OK

ac-s configuration screen

2. The lower left-hand corner of the window has a **Channel Calibration** box with control buttons for **Zero All** and **Auto Cal**.
  - The **Zero All** button sets all of the calibration offsets to 0.0 so you can collect the raw (uncorrected) values from the meter. Note that these values are always temperature-compensated.
  - The **Auto Cal** control collects and averages the new offsets for a device file. Note the option for number of samples to average. Select the number of individual samples you wish to average by using this control. The default is 10 samples.
3. Once you have selected the number of samples over which to average, you then can engage the **Auto Cal** control. A temporary pop-up window will notify you as the calibration takes place. After the operation is complete, exit the Configuration dialog box by pressing **OK** in the lower right hand corner.
4. When you exit you will be prompted as to whether you want to save the new offset values.

#### **Caution**

Unless you deliberately save a device file the offsets you have collected  
will not be permanently stored.

---

5. You may write over an older device file if you choose, but the program will ask you if you are sure you want to do this. Otherwise you may write to a new file.
6. Before you can use the new values, reload the newly made device file using **File/Open Device**.



## Appendix A—ac-9 Meter Files

### ac-9 Configuration File

Configuration files (with file extension .DEV) give calibration and other information specific to a particular unit. These files are tab-delimited text files and have the following format:

|            |   |
|------------|---|
| Line 1     | Device name.  |
| Line 2     | Serial number: the serial number of the device that was used to collect the data.   |
| Line 3     | Version number of the following structure. This should be “2.”  |
| Line 4     | Reserved for future use.  |
| Line 5     | Calibrations for depth meter: there are two values, the first is an offset and the second is a multiplier.  |
| Line 6     | RS-232 baud rate that the instrument uses   |
| Line 7     | Optical path length through water (in meters).  |
| Line 8     | Number of temperature compensation bins.  |
| Line 9     | Several values (the count is given in the preceding line); each value is the average temperature of the temperature bin.  |
| Line 10–27 | Each line describes one channel of the instrument; the first three fields are:<br>label for identifying the channel<br>color for plotting<br>clean water calibration constant (offset)<br><br>The rest of the line contains temperature compensation values that correspond to the temperature bins given in the previous line. |
| Line 28    | Reserved for future use   |
| Line 29    | Extra capabilities mask. This is a list of capabilities that the meter may have in addition to the standard product. Currently, only one such capability exists: an external temperature sensor. If the first number on this line is non-zero, the meter supports such a sensor.  |

### ac-9 WETView Data Files

WETView saves data files in tab-delimited format. The file format of WETView files is described below.

Line 1: Header line. Identifies the version of the program that created the file, and the time and date of creation.

Lines 2–30: Exactly the same format as the configuration file. These lines contain the calibration information that was used while collecting the data.

Line 31: Collection bin size. The number of samples that were averaged into each bin during collection.

#### Data lines

The rest of the file contains lines of data values. Each line of data starts with a zero-based time stamp. Times are given in milliseconds. The rest of the fields on the line

are the data from the device. The data is given as absorption values, after referencing, calibration and temperature correction.

### Reference lines

A reference line is a data line that has some extra information appended to it.

1. The first number after the data values is the temperature (deg C).
2. The second number is the sample rate (samples per second).
3. The third number is the depth (in meters, if the meter supports depth).
4. The fourth number is the external temperature, if the meter supports it.
5. The rest of the line contains channel reference values.

A full data line with references then looks like:

Time Chan<sub>1</sub> Chan<sub>2</sub> ... Chan<sub>n</sub> Temp Rate Depth ExtTemp Ref<sub>1</sub> Ref<sub>2</sub> ... Ref<sub>n</sub>

The reference values are raw values; no scaling or calibration is applied to them. These are the values that are used in computing the data values for the following set of data lines.

There is a possible ambiguity when the collection bin size is greater than one. Since a bin contains multiple samples, some of the samples in a bin may be computed with one set of reference values and others computed with a different set. The ac-9 uses one set of references for every ten samples. If the collection bin size is six, the first bin (six samples) uses the same reference values, but the second bin uses one set of references for the first four samples and a different set for the last two samples. This makes deconvoluting to the raw data rather complicated.

If you need to deconvolute to raw data, choose a bin size of 1, 2, 5, 10 or an integer multiple of ten. This ensures that the reference values given for a bin apply to all samples in that bin.

### Example

For example, given the configuration file:

```
ac9 Absorption and Attenuation Meter
04000103      ; serial number
2             ; structure version number
Reserved      ; reserved for future
5.41000 0.27000 ; depth calibration
19200        ; baud rate
0.10         ; pathlength = 10 cm
a650 Blue  1.20000 23.00000 2.43000
a560 Green  1.13000 23.00000 2.45000
a532 Brown  1.24000 23.00000 2.65000

c712 Red    1.08000 23.00000 2.31000
c676 Magenta 1.41000 23.00000 2.74000
c660 Gray   1.55000 23.00000 2.56000
a488 LtBlue 1.16000 23.00000 2.44000
a456 LtGreen 1.19000 23.00000 2.52000
a412 Yellow 1.33000 23.00000 2.54000
c650 Blue   0.98000 23.00000 2.47000
c560 Green  1.58000 23.00000 2.36000
c532 Brown  1.94000 23.00000 2.51000
a712 Red    1.32000 23.00000 2.57000
```

```

a676 Magenta 1.74000 23.00000 2.46000
a660 Gray 1.83000 23.00000 2.49000
c488 LtBlue 1.35000 23.00000 2.38000
c456 LtGreen 1.50000 23.00000 2.33000
c412 Yellow 1.11000 23.00000 2.39000
0.0035 0.004 0.015 0.02 0.015 0.02 2500 2000 1000 0.01 0.01 ...
0 ; auxiliary capabilities

```

The data file for an ac-9 begins as:

```

WetView ver 5.0 06/22/93 08:17:44
ac9 Absorption and Attenuation Meter
04000103 ; serial number
2 ; structure version number
Reserved ; reserved for future
5.41000 0.27000 ; depth calibration
19200 ; baud rate
0.10 ; pathlength = 10 cm
a650 Blue 1.20000 23.00000 2.43000
a560 Green 1.13000 23.00000 2.45000
a532 Brown 1.24000 23.00000 2.65000

c712 Red 1.08000 23.00000 2.31000
c676 Magenta 1.41000 23.00000 2.74000
c660 Gray 1.55000 23.00000 2.56000
a488 LtBlue 1.16000 23.00000 2.44000
a456 LtGreen 1.19000 23.00000 2.52000
a412 Yellow 1.33000 23.00000 2.54000
c650 Blue 0.98000 23.00000 2.47000
c560 Green 1.58000 23.00000 2.36000
c532 Brown 1.94000 23.00000 2.51000
a712 Red 1.32000 23.00000 2.57000
a676 Magenta 1.74000 23.00000 2.46000
a660 Gray 1.83000 23.00000 2.49000
c488 LtBlue 1.35000 23.00000 2.38000
c456 LtGreen 1.50000 23.00000 2.33000
c412 Yellow 1.11000 23.00000 2.39000
0.0035 0.004 0.015 0.02 0.015 0.02 2500 2000 1000 0.01 0.01 ...
0 ; auxiliary capabilities
1 ; acquisition binsize
0 1.53275 1.13881 1.31256 0.51315 1.41231 2.51512 ...
160 1.53070 1.13055 1.31583 0.51229 1.41614 2.51415 ...
330 1.53627 1.13536 1.31511 0.51467 1.41151 2.51614 ...
(etc)

```

The fields for the reference line for the ac-9 are:

| field # | contents               |
|---------|------------------------|
| 1       | time in milliseconds   |
| 2       | absorption for 650 nm  |
| 3       | absorption for 560 nm  |
| 4       | absorption for 532 nm  |
| 5       | attenuation for 712 nm |
| 6       | attenuation for 676 nm |
| 7       | attenuation for 660 nm |
| 8       | absorption for 488 nm  |
| 9       | absorption for 456 nm  |
| 10      | absorption for 412 nm  |
| 11      | attenuation for 650 nm |
| 12      | attenuation for 560 nm |

|    |                                  |
|----|----------------------------------|
| 13 | attenuation for 532 nm           |
| 14 | absorption for 712 nm            |
| 15 | absorption for 676 nm            |
| 16 | absorption for 660 nm            |
| 17 | attenuation for 488 nm           |
| 18 | attenuation for 456 nm           |
| 19 | attenuation for 412 nm           |
| 20 | temperature (°C)                 |
| 21 | sample rate (samples per second) |
| 22 | depth (meters)                   |
| 23 | reserved                         |
| 24 | reference for a650               |
| 25 | reference for a560               |
| 26 | reference for a532               |
| 27 | reference for c712               |
| 28 | reference for c676               |
| 29 | reference for c660               |
| 30 | reference for a488               |
| 31 | reference for a456               |
| 32 | reference for a412               |
| 33 | reference for c650               |
| 34 | reference for c560               |
| 35 | reference for c532               |
| 36 | reference for a712               |
| 37 | reference for a676               |
| 38 | reference for a660               |
| 39 | reference for c488               |
| 40 | reference for c456               |
| 41 | reference for c412               |

The information in a data line is exactly the same as the first 19 fields of the reference line.

### ac-9 Raw Data

Below is one record of raw data that was captured from the device. It has been converted to hexadecimal, formatted and commented. Note that the two-byte and four-byte values are byte-reversed. That is, the bytes within a word are read from right to left. This reflects the way the Intel 386 and 486 processors store values in memory.

|                                  |  |
|----------------------------------|--|
| 00FF00FF                         | registration word                              |
| 7A02                             | record length, from this word through checksum |
| 21010000                         | serial number                                  |
| 0000                             | status   |
| DB13                             | sample rate                                    |
| 1600                             | depth  |
| 34FF                             | external temperature                           |
| 6410 171E 89 2D20 0F 0319 1B ... | first sample, time and data                    |
| 7510 171E CB 2C20 EE 0119 A3 ... | second sample, time and data                   |
| 8510 161E 84 2E20 0A 0119 F2 ... | o  |
| 9510 131E F1 2920 DE FE18 DF ... | o  |
| A510 151E 9B 2B20 BC 0119 51 ... | o  |
| B510 151E 7A 2B20 BA 0119 9A ... |  |

```

C510 161E 5D 2C20 99 0019 49 ...
D510 141E F5 2B20 85 0019 B1 ...
E510 121E 3F 2820 76 FF18 11 ...
F610 141E 05 2A20 5E 0019 35 ... tenth sample, time and data
7804 C8 B604 C0 A603 66 ... 3 references and temperature
9701 temperature
65D00000 checksum (FF00FF00 through refs)
00000000 padding (4 null bytes)

```

**Registration word** = 00FF00FF hex  
**Record length** = 7A02 hex = 634 bytes  
**Serial number** = 00000121  
 This is the thirty-third ac-9 meter produced. Production numbers begin with 0100 hex.  
**Status** = 0000 Zero indicates normal operation.  
**Sample rate** = 14F9 hex = 5083  
 This is a count of the time used to take one sample. It is scaled by 0.0000316 to match the hardware's clock rate, then inverted to give samples per second:  
 $1 / (5083 * 0.0000316) = 6.226$   
**Depth** = 0016 hex = 22  
 This reflects the voltage read from the depth sensor. It is scaled and offset by the values from the configuration file.  
 $22 * 0.3 + 5.3 = 11.9$  meters  
**External Temp** = 34FF hex  
 This word is ignored for this meter because the meter does not have an external temperature probe in it. This is determined by the last line of the configuration file: the value is zero.

For the purpose of this discussion, we skip to the third-from-last line of the data record. This is the reference line, and values here will be used in the computation of the absorption values for each channel.

Reference a610 = 7804 C8 hex = 13108344  
 This is the reference value for the first channel (ab610). This is a 24-bit value indicating the output from an analog to digital converter. This represents a fraction of the input voltage. To convert from counts to this fraction, the value is divided by  $2^{24}$  (16777216).

Temperature = 13108344/16777216 = 0.7813  
 = 0F01 hex = 271  
 The temperature is given as a reading from a thermistor. The manufacturer of the thermistor provides a table correlating the reading (counts) to temperature. That table fits the polynomial equation given above. Using 407 counts, we get: 10.61831  
 $+ 0.045113 * 271$   
 $+ -4891.32 * 1/271$   
 $+ 208130.2 * 1/271^2$   
 $+ 1171473 * 1/271^3 = 7.69$  degrees C

Absorption 610 = 171E 89 hex = 8986135  
 This is the signal value for the first channel (a610). This 24-bit value is converted to raw voltage in the same manner as for reference values.  
 $8986135 / 16777216 = 0.5356$  VDC

Raw absorption is computed from the signal and reference values as described above:

$$a_{raw} = \frac{-\ln\left(\frac{E_{sig}}{E_{ref}}\right)}{Z}$$

The device's optical path length is read from the configuration file.

$$a_{raw} = \frac{-\ln\left(\frac{0.5356}{0.7813}\right)}{0.25} = 1.5103 \text{ meter}^{-1}$$

The temperature correction is then applied using the temperature from the reference line and the channel's correction table from the configuration file. The approximate correction value is linearly interpolated from the table. First, the correct temperature bin is determined by finding the two bin temperatures,  $T_0$  and  $T_1$ , that bracket the current temperature. Then, using the values,  $\Delta_{T_n}$  and  $\Delta_{T_{n+1}}$ , from the table,

$$\Delta_T = \Delta_{T_n} + \frac{(T - T_0)}{(T_1 - T_0)} * (\Delta_{T_{n+1}} - \Delta_{T_n})$$

where,

- $\Delta_T$ =compensation constant
- $T$ =current temperature, 7.69
- $T_0$ =first bin temperature, 5.5233

$T_l$ =second bin temperature, 8.4553

$\Delta_{Tn}$ =first value, 0.1411

$\Delta_{Tn+l}$ =second value, 0.1028

Using these values,

$$\Delta_T = 0.1411 + \frac{(7.69 - 5.5233)}{(8.4553 - 5.5233)} * (0.1028 - 0.1411) = 0.1127\text{m}^{-1}$$

Subtracting this from the raw absorption,

$$\begin{aligned} a' &= a_{raw} - \Delta_T \\ &= 1.5103 - 0.1127 \\ &= 1.3976 \end{aligned}$$

Finally, adding in the calibration offset for a610,

$$\begin{aligned} a &= a' + C \\ &= 1.3976 + 7.6242 \\ &= 9.0218 \end{aligned}$$

Checksum = 65D00000 = 53349

The checksum is the sum of all the bytes of the record, beginning with the registration word and ending with the temperature. For this record, the checksum is 53349. This is used to verify that the record was received correctly by WETView.

Padding = 00000000

These four null bytes separate records. They are ignored. The advantage of having them is that if a character is lost in transmission, WETView will read one byte past the end of the corrupted record. If there were no padding bytes, the first byte of the next record would have been read, corrupting that record as well. With the padding bytes, however, if a character is lost in transmission, then WETView will read one byte past the end of the record, which will be a padding byte, and it can be safely discarded. The next record will be read correctly.

## Appendix B—ac-s Meter Files

### ac-s Configuration File

ac-s configuration files (with file extension .DEV) give calibration and other information specific to a particular meter. These files are tab-delimited text files and have the following format:

|   |   |
|---|---|
| Line 1  | Device name.  |
| Line 2  | Serial number. The serial number of the device that was used to collect the data.   |
| Line 3  | Version number of the following structure. This should be “3” or greater.   |
| Line 4  | Reserved for future use.  |
| Line 5  | Calibrations for depth meter: there are two values, the first is an offset and the second is a multiplier.  |
| Line 6  | Serial port baud rate the instrument uses.  |
| Line 7  | Optical path length through water (in meters).  |
| Line 8  | Number of output wavelengths.   |
| Line 9  | Number of temperature compensation bins.  |
| Line 10   | Several values (the count is given in the preceding line); each value is the average temperature for each temperature bin.  |
| Lines 11–varies<br>(dependent on value in line 8) | <p>Each line describes one wavelength pair of the instrument; the first five fields are:</p> <ul style="list-style-type: none"> <li>label for identifying the c wavelength</li> <li>label for identifying the a wavelength</li> <li>color for plotting within WETView</li> <li>clean water calibration constant for attenuation, c</li> <li>clean water calibration constant for absorption, a</li> </ul> <p>The rest of the line contains temperature compensation values that correspond to the temperature bins given in the previous line. The first <math>n</math> values are for c, the next <math>n</math> values are for a, where <math>n</math> is the number of temperature bins.</p> |
| Varies<br>(dependent on value in line 8)          | Reserved for future use.  |



For example:

ACS Meter

53000002 ; Serial number  
3 ; structure version number

Reserved

0 1 ; Depth calibration  
115200 ; Baud rate  
0.25 ; Path length (meters)  
80-90 ; output wavelengths  
63 ; number of temperature bins

7.632727 8.171312 8.706667 9.219706 9.727826...

|        |        |     |          |          |          |          |          |             |
|--------|--------|-----|----------|----------|----------|----------|----------|-------------|
| C398.6 | A397.5 | 8   | 1.691751 | -0.80038 | 0.018293 | 0.017729 | 0.017763 | 0.016839... |
| C402.2 | A400.9 | 10  | 1.770073 | -0.41443 | 0.018396 | 0.018062 | 0.017715 | 0.016939... |
| C405.7 | A404.5 | 11  | 1.81221  | -0.09669 | 0.018499 | 0.018026 | 0.017409 | 0.016157... |
| C409.4 | A408.2 | 12  | 1.82833  | 0.132231 | 0.015788 | 0.015389 | 0.014772 | 0.013682... |
| C413.0 | A411.8 | 13  | 1.84002  | 0.285005 | 0.018766 | 0.01837  | 0.018024 | 0.016658... |
| C417.0 | A415.7 | 15  | 1.850093 | 0.382613 | 0.016909 | 0.01643  | 0.015824 | 0.015166... |
| C420.9 | A419.6 | 16  | 1.856859 | 0.453001 | 0.016226 | 0.015822 | 0.015416 | 0.014600... |
| :      |        |     |          |          |          |          |          |             |
| C730.6 | A730.7 | 106 | -0.28523 | 0.017783 | 0.004847 | 0.004845 | 0.004855 | 0.004802    |
| C734.2 | A734.4 | 107 | -0.50568 | -0.17825 | 0.004823 | 0.004771 | 0.004818 | 0.004686    |
| C737.9 | A737.9 | 109 | -0.676   | -0.32109 | 0.004798 | 0.004803 | 0.004837 | 0.004833    |
| C741.5 | A741.4 | 110 | -0.78822 | -0.40855 | 0.004683 | 0.004676 | 0.004701 | 0.004667    |
| C745.2 | A744.9 | 111 | -0.85581 | -0.45682 | 0.004321 | 0.00431  | 0.004344 | 0.004257    |
| C748.8 | A748.6 | 112 | -0.89734 | -0.47902 | 0.004134 | 0.004146 | 0.004179 | 0.004176    |
| C752.3 | A752.0 | 114 | -0.91907 | -0.48634 | 0.004145 | 0.004146 | 0.004201 | 0.004071    |
| C755.5 | A755.5 | 115 | -0.93167 | -0.48493 | 0.003956 | 0.00395  | 0.003961 | 0.003925    |
| 0      | 0      | 0   | 0        | 0        | 0        | 0        | 0        | 0           |

### ac-s WETView Data Files

WETView saves data files in tab-delimited format. The file format for data files is as follows:

|                     |   |
|---------------------|---|
| Line 1              | Header line. Identifies the version of the program that created the file, and the time and date of creation.  |
| Line 2–98           | Exactly the same format as the configuration file. These lines contain the calibration information that was used while collecting the data.   |
| Line 99             | Acquisition bin size  |
| Line 100            | Wavelength labels   |
| Lines 101– <i>n</i> | Data lines. Each line starts with a zero-based time stamp. Times are in milliseconds from the time the data collection starts. The first 86 columns are “c” values, followed by the “a” values. then internal temperature, , , , and. |

After the data, there are eight additional columns:

1. instrument temperature (deg C)
4. filter wheel speed diagnostic
5. pressure
6. external temperature
7. four columns of dark values

A data line then looks like:

Time  $c(\lambda_1), c(\lambda_2), \dots, c(\lambda_n), a(\lambda_1), a(\lambda_2), \dots, a(\lambda_n)$  int. temp, diagnostic, pressure, ext. temp, dark, dark, dark, dark

Example data file:

```

WetView
ver 7.0      11/3/03   9:26:13
ACS Meter
53000002           ; serial number
3                 ; structure version number
Reserved
0                 1           ; Depth calibration
115200            ; Baud rate
0.25              ; Path length (meters)
80–90             ; Output wavelengths
33                ; number of temperature bins
                                0 ; temperature bins

C399.574 A397.577      8 0.834057 -0.40889      0.012469 0.012711  0.01349...
C403.161 A401.155     10 0.839778 -0.18519      0.012172 0.012505  0.012554...
C406.76  A404.773     11 0.849354 0.051999      0.011779 0.01243   0.012282...
C410.482 A408.416     12 0.855933 0.245707      0.011866 0.012792  0.012571...

```

---

```

C414.072 A412.172      13 0.867732 0.384503      0.013137 0.013593 0.01367...
C417.901 A416.138      15 0.881948 0.478344      0.01188 0.012485 0.012429...
C421.964 A420.024      16 0.902806 0.541544      0.012572 0.013131 0.013225...
C426.069 A424.002      17 0.92668 0.586729      0.013561 0.014083 0.01365...
C430.248 A428.324      18 0.944917 0.625062      0.01167 0.012394 0.01234...
C434.416 A432.434      20 0.960641 0.661356      0.012023 0.012732 0.012885...
C438.594 A436.657      21 0.973538 0.695719      0.01173 0.012489 0.012435...
C442.707 A440.827      22 0.986726 0.730411      0.011702 0.012497 0.012462...
C446.933 A445.194      23 0.997492 0.764348      0.011456 0.01212 0.012393...
C451.282 A449.474 LtBlue 1.009458 0.798393      0.012098 0.012729 0.012767...
C455.74 A453.843      26 1.023926 0.831253      0.012313 0.012988 0.013176...
o
o
o

C750.235 A748.13      112 1.201872 2.112206      0.00529 0.006012 0.006551...
C753.62 A751.57      114 1.201639 2.122247      0.005119 0.005896 0.006448...
C756.521 A754.65      115 1.200377 2.132103      0.005052 0.005757 0.006322...
0 0 0 0 0 0 0 0 0 0...
10 ; acquisition binsize ...
C399.574 C403.161 C406.76 C410.482 C414.072 C417.901 C421.964 C426.069...
4500 -0.00283 -0.00193 -0.00485 -0.00604 -0.00795 -0.00449 -0.00593 -0.00623...
9500 -0.00425 -0.00287 -0.00725 -0.00765 -0.00751 -0.00633 -0.00714 -0.00646...
14500 -0.00554 -0.00699 -0.00338 -0.00722 -0.00529 -0.0079 -0.00491 -0.00707...
19500 -0.00561 -0.00444 -0.00816 -0.0049 -0.00278 -0.0085 -0.00445 -0.00557...
24500 -0.00364 -0.00055 -0.0018 -0.00276 -0.00429 -0.00345 -0.00506 -0.0055...
29500 -0.0036 -0.00869 -0.00621 -0.0041 -0.00558 -0.00591 -0.00664 -0.0048...
34500 0.00088 -0.00263 -0.00256 -0.00676 -0.00609 -0.0044 -0.00618 -0.00556...
39500 0.00213 -0.0038 -0.00233 -0.00397 -0.00695 -0.00422 -0.00369 -0.00527...
o
o
o

```

### ac-s Raw Data

Below is one record of raw data that was captured from the device. It has been converted to hexadecimal, formatted and commented.

|          |   |
|----------|---|
| ff00ff00 | Packet registration   |
| 02D0     | Record length of full packet (not including chksum) = 720         |
| 04       | Packet Type 03 and above designates an ac-s meter                 |
| 01       | unused by ac-s  |
| 53       | meter type—53 indicates an ac-s                                   |
| 000002   | serial number   |
| 4e1a     | sample to sample delta time in 25 $\mu$ s counts (0.49985 sec)    |
| 01ba     | Reserved  |
| 02a1     | Reserved  |
| 7ae4     | raw external temp counts = 26.82 $^{\circ}$ C                     |
| b9d7     | raw internal temp counts = 27.93 $^{\circ}$ C                     |
| 01d5     | Reserved  |
| 02b0     | Reserved  |
| 00071b02 | time in milliseconds since power up (465666 = 7.761 mins)         |
| 01       | Reserved  |
| 56       | Number of output wavelengths = 80–90 decimal (86 in this example) |
| 0405     | raw cref <sub>1</sub> counts                                      |
| 0363     | raw aref <sub>1</sub> counts                                      |
| 04f4     | raw csig <sub>1</sub> counts                                      |
| 0310     | raw asig <sub>1</sub> counts                                      |
| 0498     | raw cref <sub>2</sub> counts                                      |
|          |   |
| 20bb     | raw cref <sub>86</sub> counts                                     |
| 19bf     | raw aref <sub>86</sub> counts                                     |
| 2c49     | raw csig <sub>86</sub> counts                                     |
| 2c1c     | raw asig <sub>86</sub> counts                                     |
| 2243     | Checksum  |
| 00       | pad byte  |

|                      |  |
|----------------------|--|
| Registration word    | = FF00FF00 hex   |
| Record length        | = 02D0 hex = 720 bytes   |
| Packet type          | = 04<br>Reserved for future use.   |
| Reserved             | = 01   |
| Meter Type           | = 53   |
| Serial Number        | = 000002<br>This is the second ac-s meter produced (production numbers begin with 0000 hex).   |
| Status               | = 4E1A Sample to sample delta time in 25 us counts (0.49985 sec)   |
| Reserved             | <2 undefined bytes><br>Reserved for future use.  |
| Reserved             | <2 undefined bytes><br>Reserved for future use.  |
| External Temperature | = 7AE4 hex = 31460<br>The external temperature is derived from a voltage drop across a thermistor. The temperature is calculated from a polynomial to convert counts to temperature.<br>The polynomial to convert counts to temperature is:<br>$y = a \times \text{counts}^3 + b \times \text{counts}^2 + c \times \text{counts} + d$ where $a = -7.1023317 \text{ E-13}$ , $b = 7.09341920 \text{ E-8}$ , $c = -3.87065673 \text{ E-3}$ , and $d = 95.8241397$<br>which yields an external temperature of 22.14°C   |
| Internal Temperature | = B9D7 hex = 47575<br>The internal temperature is derived from a voltage drop across a thermistor. The temperature is calculated from an equation to convert the thermistor resistance to temperature. The equation for converting the raw A/D counts to volts is:<br>$\text{volts} = (5 \times \text{counts} / 65535)$ which yields 3.6297<br>The volts are then converted to resistance using the equation:<br>$\text{resistance} = 10,000 \times \text{volts} / (4.516 - \text{volts})$ which yields 40,956.7 ohms<br>The equation to convert resistance to temperature is: |

$$temp = \frac{1}{a + b \times \ln(res) + c \times (\ln(res))^3} - 273.15$$

where a = 0.00093135, b = 0.000221631, and c = 0.000000125741

which yields

$$T = \frac{1}{a + b \ln 40956 + c(\ln 40956)^3} - 273.15 = 17.91 \text{ } ^\circ\text{C}$$

Reserved <4 undefined bytes>  
Reserved for future use.

Time since power up = 0071b02 hex = 465666ms = 7.761 minutes

Number of output wavelengths = 56 hex = 86

Data, wavelength 1 = 0405 0363 04f4 0310 hex = 1029 867 1268 784

This is the data read for the first wavelength: c reference, a reference c, and a channels, respectively. Raw absorption is computed from the signal and reference values as described above:

$$a_{raw} = \frac{-\ln\left(\frac{E_{sig}}{E_{ref}}\right)}{Z}$$

Which yields  $a_{raw} = -\ln(784/867)/0.25 = 0.402 \text{ m}^{-1}$  and  $c_{raw} = -\ln(1268/1029)/0.25 = -0.835 \text{ m}^{-1}$ .

The channel's calibration constant and the device's optical path length are read from the configuration file.

The temperature correction is then applied using the internal temperature and the channel's correction table from the configuration file. The approximate correction value is linearly interpolated from the table. First, the correct temperature bin is determined by finding the two bin temperatures,  $T_0$  and  $T_1$ , that bracket the current temperature. Then, using the values,  $\Delta_{Tn}$  and  $\Delta_{Tn+1}$ , from the table,

$$K_T = v_0 + \frac{(T - T_0)}{(T_1 - T_0)} * (v_1 - v_0)$$

where,

$\Delta_T$  = compensation constant

$T$  = current temperature, 27.93

$T_0$  = first bin temperature, 27.75

$T_1$  = second bin temperature, 28.2625

$\Delta_{Tn}$  = first value, 0.010124

$\Delta_{Tn+1}$  = second value, 0.015096

Using these values,

$$\begin{aligned} \Delta_T &= 0.010124 + (27.93 - 27.75) / (28.2625 - 27.75) * (0.015096 - 0.010124) \\ &= 0.012 \end{aligned}$$

Subtracting this from the raw absorption,

$$a' = 0.402 - 0.012 = 0.039$$

Finally, adding in the calibration offset for a397.5,

$$a = a' + C = 0.039 - 0.431 = -0.392$$

Checksum = 2243 = 8771

The checksum is the sum of all the bytes of the record, beginning with the registration word and ending with the byte just before this checksum. For this record, the checksum is 8771. This is used to verify that the record was received correctly by WETView.



### Revision History

| <b>Revision</b> | <b>Date</b> | <b>Revision Description</b>                                   | <b>Originator</b>      |
|-----------------|-------------|---|------------------------|
| A               | 02/08/00    | Begin revision control  | H. Van Zee             |
| B               | 04/12/00    | Correct calibration offset for ac-9 (DCR 20)                  | C. de Lespinasse       |
| C               | 05/23/00    | Correct data and device files for HiStar (DCR 31)             | D. Hankins, H. Van Zee |
| D               | 07/12/00    | Correct equation on temperature correction algorithm (DCR 47) | C. de Lespinasse       |
| E               | 4/12/01     | Delete references to three-spectrometer HiStar (DCR 102)      | D. Hankins             |
| E1              | 2/4/04      | Update to include ac-s, delete HiStar references              | B. Rhoades             |
| F               | 2/9/04      | Approved revisions in E1 (DCR 359)                            | D. Hankins, B. Rhoades |
|                 |             |   |                        |
|                 |             |   |                        |