

PO Box 518
620 Applegate St.
Philomath, OR 97370



(541) 929-5650
Fax (541) 929-5277
www.wetlabs.com

ac-s Calibration

Date	9/20/2016	Customer	U of Maine	RO #	31754
S/N #	acs007	Technician	cad	Cal #	15

Wavelength range: 400 to 730 nm

Tcal: 20.9 deg C

Ical: 26.3 deg C

Acal: 28.7 deg C

Temperature range: 7.6 to 45.0 deg C

Zero: 25.5 deg C

Definitions

Wavelength range: Wavelengths for which instrument use is recommended.

Tcal: Water temperature in flow tubes during water calibration.

Ical: Internal instrument temperature during water calibration.

Acal: Internal instrument temperature when recording air offsets in the *.cal file.

Temperature range: Recommended internal instrument operating range.

Zero: Internal instrument temperature at which the temperature correction is zero in the device file.