

PO Box 518  
620 Applegate St.  
Philomath, OR 97370



(541) 929-5650  
Fax (541) 929-5277  
[www.wetlabs.com](http://www.wetlabs.com)

## ac-s Calibration

Date 9/21/2016

S/N # acs279

Cal # 1

---

Wavelength range: 400 to 730 nm

Tcal: 21.3 deg C

Ical: 18.7 deg C

Acal: 23.8 deg C

Temperature range: 0.3 to 35.0 deg C

Zero: 25.5 deg C

---

### Definitions

Wavelength range: Wavelengths for which instrument use is recommended.

Tcal: Water temperature in flow tubes during water calibration.

Ical: Internal instrument temperature during water calibration.

Acal: Internal instrument temperature when recording air offsets in the \*.cal file.

Temperature range: Recommended internal instrument operating range.

Zero: Internal instrument temperature at which the temperature correction is zero in the device file.