



# Sea-Bird Electronics, Inc.

13431 NE 20<sup>th</sup> St. Bellevue, Washington 98005 USA  
www.seabird.com

Phone: (425) 643-9866  
Fax: (425) 643-9954  
Email: seabird@seabird.com

## Pressure Test Certificate

Test Date: 2016-12-14

Description: SBE-19Plus SeaCat Profiler

### Sensor Information:

Replaced the main piston "O"-Rings.

Model Number: SBE-19P

Serial Number: 4547

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For: 15 Minutes

High Pressure Test: 870 PSI Held For: 15 Minutes

Passed Test: True

Tested By: nd

High pressure is generally equal to the maximum depth rating of the instrument

