



**APPLICATION NOTE NO. 11 General**

**Revised February 2011**

**PAR Light Sensors**

Sea-Bird has several application notes dealing with PAR sensors from various manufacturers; this application note provides an overview of PAR measurements and units, and is applicable to all PAR sensors.

**PAR** is an abbreviation for **Photosynthetically Available Radiation** (also called **Photosynthetically Active Radiation**). Solar radiation reaching Earth’s surface is a mixture of ultraviolet light, visible light, and near-visible infrared radiation. All of this radiation conveys heat; the portion between approximately 400 and 700 nm wavelength can be captured and used by photo-autotrophs (organisms capable of obtaining energy directly from sunlight), and is called PAR.

Irradiance is the flux of solar radiation incident on a surface per unit time per unit area and is reported in units of energy content (Watts/m<sup>2</sup>) or photon content (quanta/m<sup>2</sup> sec or μEinsteins/m<sup>2</sup> sec). Conversion from energy to photon content can be made with Planck’s equation, provided that the light wavelength is known. The energy of a photon is related to its wavelength as follows:

$$E = hc / \lambda$$

where

h = Planck’s constant (6.626 x 10<sup>-34</sup> Joules sec)  
c = speed of light (2.998 x 10<sup>8</sup> m/sec)  
λ = wavelength (m)

This equation provides the energy for a single wavelength. For a broad spectrum PAR sensor, a wavelength of approximately 550 nm (550 x 10<sup>-9</sup> m) is typically used for the conversion.

“For marine atmospheres with sun altitudes above 22 degrees, the quanta/watt ratio for the region 400 to 700 nm is 2.77 x 10<sup>18</sup> quanta/sec/Watt to an accuracy of plus or minus a few percent.” This quote and further discussion of the relationship of quanta to Watts in the water column is found in Smith and Morel (1974) Limnol. Oceanogr. 19(4):591-600.

$$E \text{ (at 550 nm)} = hc / \lambda = (6.626 \times 10^{-34} \text{ Joules sec}) * (2.998 \times 10^8 \text{ m/sec}) / (550 \times 10^{-9} \text{ m}) = 3.61 \times 10^{-19} \text{ Joules}$$

(Note: 1 / 3.61 x 10<sup>-19</sup> = 2.77 x 10<sup>18</sup> quanta/sec/Watt, the value quoted in the above reference.)

Application notes for underwater PAR sensors (**11Chelsea, 11Licor, 11QSP-L, and 11QSP-PD**) and surface PAR sensors (**11S and 47**) describe how to enter coefficients from the manufacturer’s calibration in the CTD configuration (.con or .xmlcon) file to provide SEASOFT output in μEinsteins/m<sup>2</sup>·sec. To calculate irradiance in other units:

| To convert to:                  | For Underwater PAR Sensors,<br>set Multiplier to:   | For Surface PAR Sensors,<br>multiply calculated Conversion factor by: |
|---------------------------------|---|---|
| μEinsteins/m <sup>2</sup> ·sec  | <b>1.0</b>  |   |
| μEinsteins/cm <sup>2</sup> ·sec | (1.0) / (100 cm/m) <sup>2</sup> = <b>1 x 10<sup>-4</sup></b>  |   |
| Einsteins/m <sup>2</sup> ·sec   | (1.0) / (1 x 10 <sup>6</sup> μEinsteins/Einstein) = <b>1 x 10<sup>-6</sup></b>                        |   |
| Einsteins/cm <sup>2</sup> ·sec  | (1 x 10 <sup>-6</sup> ) / (100 cm/m) <sup>2</sup> = <b>1 x 10<sup>-10</sup></b>                       |   |
| quanta/m <sup>2</sup> ·sec      | (1 x 10 <sup>-6</sup> ) * (6.022 x 10 <sup>23</sup> quanta/Einstein) = <b>6.022 x 10<sup>17</sup></b> |   |
| quanta/cm <sup>2</sup> ·sec     | (6.022 x 10 <sup>17</sup> ) / (100 cm/m) <sup>2</sup> = <b>6.022 x 10<sup>13</sup></b>                |   |
| Watts/m <sup>2</sup>            | (6.022 x 10 <sup>17</sup> ) / (2.77 x 10 <sup>18</sup> quanta/sec/Watt) = <b>0.2174</b>               |   |
| Watts/cm <sup>2</sup>           | (0.2174) / (100 cm/m) <sup>2</sup> = <b>2.174 x 10<sup>-5</sup></b>                                   |   |
| μWatts/m <sup>2</sup>           | (0.2174) * (1 x 10 <sup>6</sup> μWatts/Watt) = <b>2.174 x 10<sup>5</sup></b>                          |   |

Note: 1 Einstein = 1 mole (6.022 x 10<sup>23</sup>) of photons      1 Watt = 2.77 x 10<sup>18</sup> quanta/sec

**Notes:**

- In our SEASOFT V2 suite of programs, edit the CTD configuration (.con or .xmlcon) file using the Configure Inputs menu in Seasave V7 (real-time data acquisition software) or the Configure menu in SBE Data Processing (data processing software).
- Multiplier can also be used to scale output for comparing the shape of data sets taken at disparate light levels. For example, a multiplier of 10 would make a 10 μEinsteins/m<sup>2</sup>·sec light level plot as 100 μEinsteins/m<sup>2</sup>·sec.

---

## Application Note Revision History

| Date          | Description  |
|---------------|--|
| -             | Initial release.   |
| May 2007      | <ul style="list-style-type: none"><li data-bbox="391 331 691 363">• Incorporate Seasave V7.</li><li data-bbox="391 363 837 394">• Eliminate discussion of DOS software.</li></ul>  |
| February 2010 | <ul style="list-style-type: none"><li data-bbox="391 399 837 430">• Change Seasoft-Win32 to Seasoft V2.</li><li data-bbox="391 430 938 462">• Add information on .xmlcon configuration file.</li><li data-bbox="391 462 610 491">• Update address.</li></ul> |
| February 2011 | Correct units for $h = \text{Planck's constant}$ ( $6.626 \times 10^{-34}$ <b>Joules/sec</b> corrected to <b>Joules sec</b> )  |