



**SEA-BIRD ELECTRONICS, INC.**

1808 136th Place N.E., Bellevue, Washington 98005 USA

Telephone: (425) 643-9866 Telex: 292915 SBEI UR

Fax: (425) 643-9954

email: seabird@seabird.com

DATE	SYM	REVISION RECORD	AUTH	DR	CHK

SBE P/N 50002

**SBE 13-01 O-Ring Spares Kit**

*KIT CONTENTS*

SBE P/N	Manufacturer P/N	Primary SBE Application	QTY
24019	Sensor Medics 868080	Beckman Oxygen Oil Seal Screw	1
30072	Parker 2-017N674-70	Bulkhead Connector	1
30074	Parker 2-022N674-70	Beckman Oxygen Sensor Seal	1
30081	Parker 18mm x 4mm x 70	Beckman Oxygen Pressure Balance	1
30082	Parker 2-213N674-70	Beckman Oxygen Receptacle - End Cap	1
30816	L-Seal, LS-021	Sensor End Cap - Housing	1
231004	Backup for L-Seal, LS-021	Sensor End Cap - Housing	1
30439	Bellofram 3-50-38-BBJ	Beckman Oxygen Pressure Diaphragm	1

*O-RING USAGE:*

SBE Application	Manufacturer P/N	SBE P/N	QTY
Oil Seal Screw	Sensor Medics 868080	24019	1
Bulkhead Connectors	Parker 2-017N674-70	30072	1
Oxygen Sensor	Parker 2-022N674-70	30074	1
Pressure Balance	Parker 18mm x 4mm x 70	30081	1
Oxygen Receptacle - End Cap	Parker 2-213N674-70	30082	1
End Cap - Housing	Parker 2-220N674-70	30084	1
DO Pressure Diaphragm	Bellofram 3-50-38-BBJ	30439	1

**Sea-Bird Electronics Procedure**

PROCEDURE NUMBER: **67064**

TITLE: **SBE PN 50002 SBE 13-01 O-Ring Spares Kit**

REVISION:

EFFECTIVE DATE: **11/20/2002**

CHECKED BY: \_\_\_\_\_ PAGE 1 of 1