



Volume Scattering Function Calibration Sheet

Date 3/18/2003 Customer Emmanuel Boss
Wave: 440nm S/N# vsf-005b Tech K.C.

1. From the design, we compute the weighting function for each scattering angle for each light source by numerical interaction of the sample volume elements.

2. We determine the volume scattering functions for the calibration particles by integrating the MIE scattering functions over the size distribution.

3. We compute the portion of the total scattering coefficient for the calibration particles represented by the angular weighting functions. We then divide this by the solid angle in steradians represented by the weighting factors to get $\beta(\theta)/b$ as shown below:

$\beta(\theta)/b$ for 1.96 Micron Diameter Beads

Angle Centroid	Wavelength		
	440	530	650
100	0.00786	0.00500	0.00287
125	0.01079	0.00528	0.00257
150	0.00864	0.00354	0.00167

4. For each instrument, we obtain the scattering signals and the attenuation coefficients for a dilution series of 2-micron beads. The results for this instrument follow:

	Cp440	Counts		
		100	125	150
Dark counts		123.563	128.063	166.094
Clean water	0.014	129.419	167.065	150.774
	0.168	149.133	177.733	214.367
	0.263	162.167	202.333	252.667
	0.395	180.000	236.563	304.938
	0.745	231.516	333.161	457.968
	1.129	283.633	438.433	617.567
	1.510	337.968	540.065	768.839
	1.904	393.906	646.000	929.094

5. It is assumed that Cp is equal to b for the calibration particles. We obtain b/counts by linear regression.

	100		150	
slope, intercept	0.007134	-0.898658	0.002404	-0.346900
	4.24487E-05	0.00943685	1.43E-05	0.00639492
r	0.999823	0.00801421	0.999824	0.00798946
	125			
slope, intercept	0.003813	-0.538067		
	0.000128031	0.04194394		
r	0.994396	0.045098		

6. Multiplying b/counts by $\beta(\theta)/b$ yields:

	$\beta(\theta)/counts$
100	5.607E-05
125	4.115E-05
150	2.077E-05

7. Dividing intercept by slope yields

	Dark Counts
	123.563
	128.063
	166.094

To obtain $\beta(\theta)$, subtract the zero offset from the counts obtained, then multiply by $\beta(\theta)/counts$ from the table above.

8. The test also provides a measure of the inherent noise level of the instrument from the short term RMS deviation from the average number of counts. We translate this into the resolution of $\beta(\theta)$ (minimum detectable signal change) in units of 1/m. steradians).

	Noise (counts)	Resolution (1/[m. ster])
100	1.7964	1.01E-04
125	1.9006	7.82E-05
150	1.9268	4.00E-05

Volume Scattering Function Calibration Sheet

Date: 4/25/02

Serial # **VSF-006G**

Wavelength: 530 nm

Tester: Dan

Test Description

- From the design, we compute the weighting function for each scattering angle for each light source by numerical integration of the sample volume elements.
- We determine the volume scattering functions for the calibration particles by integrating the MIE scattering functions over the size distribution.
- We compute the portion of the total scattering coefficient for the calibration particles represented by the angular weighting functions. We then divide this by the solid angle in steradians represented by the weighting factors to get $\beta(\theta) / b$ as shown below:

$\beta(\theta) / b$ for 1.96 Micron Diameter Beads

Angle Centroid	Wavelength		
	440	530	650
100	0.00786	0.00500	0.00287
125	0.01079	0.00528	0.00257
150	0.00864	0.00354	0.00167

- For each instrument, we obtain the scattering signals and the attenuation coefficients for a dilution series of 2 micron beads. The results for this instrument follow:

	Cp530	Counts		
		100	125	150
Dark Counts		128.600	127.716	132.900
Clean water	0.004	133.250	135.233	150.516
	1.202	326.950	364.166	524.433
	2.285	502.766	570.966	856.516
	3.397	682.050	780.750	1193.483
	4.507	861.516	990.500	1525.683

- It is assumed that Cp is equal to b for the calibration particles. We obtain b/counts by linear regression.

	100		150	
slope, intercept	0.006183	-0.820484	0.0032752	-0.505311
	3.43245E-06	0.001932	1.581E-05	0.015464
r	0.999999	0.001967	0.9999301	0.0171
	125			
slope, intercept	0.005266	-0.713561		
	9.47168E-06	0.006091		
r	0.999990	0.006372		

<p>6. Multiplying b/counts by $\beta(\theta) / b$ yields:</p> <table border="1" style="margin-left: 20px;"> <thead> <tr> <th></th> <th>$\beta(\theta) / \text{counts}$</th> </tr> </thead> <tbody> <tr> <td>100</td> <td>3.091E-05</td> </tr> <tr> <td>125</td> <td>2.780E-05</td> </tr> <tr> <td>150</td> <td>1.159E-05</td> </tr> </tbody> </table> <p>To obtain $\beta(\theta)$, subtract the zero offset from the counts obtained, then multiply by $\beta(\theta) / \text{counts}$ from the table just above.</p>		$\beta(\theta) / \text{counts}$	100	3.091E-05	125	2.780E-05	150	1.159E-05	<p>7. Dividing intercept by slope yields</p> <table border="1" style="margin-left: 20px;"> <thead> <tr> <th>Dark Counts</th> </tr> </thead> <tbody> <tr> <td>128.600</td> </tr> <tr> <td>127.716</td> </tr> <tr> <td>132.900</td> </tr> </tbody> </table>	Dark Counts	128.600	127.716	132.900
	$\beta(\theta) / \text{counts}$												
100	3.091E-05												
125	2.780E-05												
150	1.159E-05												
Dark Counts													
128.600													
127.716													
132.900													

- The test also provides a measure of the inherent noise level of the instrument from the short term RMS deviation from the average number of counts. We translate this into the resolution of $\beta(\theta)$ (minimum detectable signal change) in units of 1/(m. steradians).

	Noise (counts)	Resolution (1/(m. ster))
100	0.9014	2.79E-05
125	0.8105	2.25E-05
150	0.7718	8.95E-06

Volume Scattering Function Calibration Sheet

Date: 12/19/00 Serial # VSF004R Wavelength: 650 nm Tester: Christian

Test Description

1. From the design, we compute the weighting function for each scattering angle for each light source by numerical integration of the sample volume elements.
2. We determine the volume scattering functions for the calibration particles by integrating the MIE scattering functions over the size distribution.
3. We compute the portion of the total scattering coefficient for the calibration particles represented by the angular weighting functions. We then divide this by the solid angle in steradians represented by the weighting factors to get $\beta(\theta) / b$ as shown below:

$\beta(\theta) / b$ for 2 Micron Diameter Beads

Angle Centroid	Wavelength		
	440	530	650
100	0.0076	0.0054	0.0030
125	0.0115	0.0060	0.0028
150	0.0095	0.0040	0.0017

4. For each instrument, we obtain the scattering signals and the attenuation coefficients for a dilution series of 2 micron beads. The results for this instrument follow:

	Cp650	Counts		
		100	125	150
Air		135.1867	138.7619	145.5714
Clean water	0.008	170.1636	165.8909	189.6909
	1.667	483.5082	425.3934	603.5738
	3.669	888.5667	758.3833	1168.383
	5.047	1171.517	986.5	1520.767
	6.54	1494.783	1241.717	1950.55

5. It is assumed that Cp is equal to b for the calibration particles. We obtain b/counts by linear regression.

	100		150	
slope, intercept	0.004930	-0.763034	0.0037029	-0.637293
	6.68477E-05	0.064529	4.741E-05	0.059527
r	0.999449	0.070642	0.9995085	0.0667
	125			
slope, intercept	0.006063	-0.952547		
	5.14423E-05	0.041783		
r	0.999784	0.044205		

6. Multiplying b/counts by $\beta(\theta) / b$ yields:		7. Dividing intercept by slope yields the zero offset	
	$\beta(\theta) / \text{counts}$		
100	1.479E-05		155
125	1.698E-05		157
150	6.295E-06		172

To obtain $\beta(\theta)$, subtract the zero offset from the counts obtained, then multiply by $\beta(\theta) / \text{counts}$ from the table just above.

8. The test also provides a measure of the inherent noise level of the instrument from the short term RMS deviation from the average number of counts. We translate this into the resolution of $\beta(\theta)$ (minimum detectable signal change) in units of 1/(m. steradians).

	Noise (counts)	Resolution (1/(m. ster))
100	0.889	1.31E-05
125	0.854	1.45E-05
150	0.833	5.24E-06